

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012619\
 Data File : VX007237.D
 Acq On : 25 Jan 2019 18:55
 Operator : JC/SP
 Sample : PB116609ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116609ZHE#03

Quant Time: Jan 28 01:02:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	532928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	825096	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	804969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	382616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	286017	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
35) Dibromofluoromethane	5.51	113	258821	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	8.71	98	1010003	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
62) 4-Bromofluorobenzene	11.13	95	356365	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
Target Compounds						
16) Acetone	2.45	43	13924	5.057	ug/l #	87
18) Methyl Acetate	2.78	43	26451	3.317	ug/l	97
20) Methylene Chloride	2.86	84	14330	2.195	ug/l	94
43) Isopropyl Acetate	6.46	43	29282	2.414	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012619\
Data File : VX007237.D
Acq On : 25 Jan 2019 18:55
Operator : JC/SP
Sample : PB116609ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116609ZHE#03

Quant Time: Jan 28 01:02:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

