

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012619\
 Data File : VX007246.D
 Acq On : 25 Jan 2019 22:55
 Operator : JC/SP
 Sample : K1015-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-012419-B

Quant Time: Jan 28 01:30:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	527492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	821355	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	793525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	378942	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289524	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
35) Dibromofluoromethane	5.51	113	257174	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	8.71	98	998260	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
62) 4-Bromofluorobenzene	11.13	95	352602	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	74587	27.368	ug/l	91
18) Methyl Acetate	2.78	43	24843	3.148	ug/l	97
20) Methylene Chloride	2.86	84	689216	133.911	ug/l	95
25) 2-Butanone	4.70	43	15461	3.965	ug/l #	82
29) Tetrahydrofuran	5.16	42	15993	6.311	ug/l	91
43) Isopropyl Acetate	6.46	43	27289	2.260	ug/l	97
95) Naphthalene	13.83	128	1412277	51.727	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012619\
Data File : VX007246.D
Acq On : 25 Jan 2019 22:55
Operator : JC/SP
Sample : K1015-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-04-012419-B

Quant Time: Jan 28 01:30:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

