

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012619\
 Data File : VX007254.D
 Acq On : 26 Jan 2019 02:27
 Operator : JC/SP
 Sample : K1238-50
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S5

Quant Time: Jan 28 02:32:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	533130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	831121	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	799361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	385700	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	288022	51.69	ug/l	0.00
Spiked Amount			Recovery	=	103.38%	
35) Dibromofluoromethane	5.51	113	256280	48.15	ug/l	0.00
Spiked Amount			Recovery	=	96.30%	
50) Toluene-d8	8.71	98	1012429	51.97	ug/l	0.00
Spiked Amount			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.13	95	361205	53.18	ug/l	0.00
Spiked Amount			Recovery	=	106.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	69839	25.355	ug/l	98
18) Methyl Acetate	2.78	43	13670	1.714	ug/l #	89
20) Methylene Chloride	2.86	84	945963	182.057	ug/l	94
25) 2-Butanone	4.70	43	15392	3.905	ug/l	94
29) Tetrahydrofuran	5.16	42	16340	6.380	ug/l	97
43) Isopropyl Acetate	6.46	43	24008	1.965	ug/l	94
95) Naphthalene	13.83	128	44030	1.584	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012619\
Data File : VX007254.D
Acq On : 26 Jan 2019 02:27
Operator : JC/SP
Sample : K1238-50
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S5

Quant Time: Jan 28 02:32:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

