

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012619\
 Data File : VX007238.D
 Acq On : 25 Jan 2019 19:21
 Operator : JC/SP
 Sample : PB116609ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB116609ZHE#04

Quant Time: Jan 28 01:04:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	532098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	820632	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	800654	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	384985	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	284173	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
35) Dibromofluoromethane	5.50	113	256489	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.71	98	1006188	52.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.62%	
62) 4-Bromofluorobenzene	11.13	95	351890	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
Target Compounds						
16) Acetone	2.45	43	14884	5.414	ug/l #	87
18) Methyl Acetate	2.78	43	25636	3.220	ug/l	94
20) Methylene Chloride	2.86	84	13632	2.065	ug/l	96
43) Isopropyl Acetate	6.46	43	27038	2.241	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012619\
Data File : VX007238.D
Acq On : 25 Jan 2019 19:21
Operator : JC/SP
Sample : PB116609ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB116609ZHE#04

Quant Time: Jan 28 01:04:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21

Instrument :

MSVOA_X

ClientSampleId :


PB116609ZHE#04

Tot Ion:168 Resp: 532098

Ion Ratio Lower Upper

168 100

99 50.1 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

200000



150000

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#16

Acetone

Concen: 5.414 ug/l

RT: 2.45 min Scan# 212

Delta R.T. -0.00 min

Lab File: VX007238.D


Acq: 25 Jan 2019 19:21

Tgt Ion: 43 Resp: 14884

Ion Ratio Lower Upper

43 100

58 23.9 25.0 37.6#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.45

10000

8000


6000


4000

2000

0

Time--> 2.40 2.45 2.50









#18
Methyl Acetate
Concen: 3.220 ug/l
RT: 2.78 min Scan# 267
Delta R.T. -0.00 min
Lab File: VX007238.D
Acq: 25 Jan 2019 19:21




Instrument :
MSVOA_X
ClientSampleId :
PB116609ZHE#04


Tot Ion: 43 Resp: 25636
Ion Ratio Lower Upper
43 100
74 26.1 18.6 28.0



#20
Methylene Chloride
Concen: 2.065 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.00 min
Lab File: VX007238.D
Acq: 25 Jan 2019 19:21

Tgt Ion: 84 Resp: 13632
Ion Ratio Lower Upper
84 100
49 112.5 91.8 137.6
51 30.2 28.5 42.7
86 58.3 50.2 75.2





#33

1,2-Dichloroethane-d4

Concen: 51.102 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21

Instrument :

MSVOA_X

ClientSampleId :


PB116609ZHE#04

Tot Ion: 65 Resp: 284173

Ion Ratio Lower Upper

65 100

67 54.2 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

120000

100000

80000



60000

40000

20000

0

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21


Tgt Ion: 114 Resp: 820632

Ion Ratio Lower Upper

114 100

63 17.8 0.0 34.8

88 14.8 0.0 28.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0

6.86

400000


300000


200000

100000

0

Time-->





#35

Dibromofluoromethane

Concen: 48.806 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21

Instrument :

MSVOA_X

ClientSampleId :

PB116609ZHE#04


Tot Ion:113 Resp: 256489

Ion Ratio Lower Upper

113 100

111 102.3 81.6 122.4




192 22.7 16.7 25.1



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#43

Isopropyl Acetate

Concen: 2.241 ug/l

RT: 6.46 min Scan# 870

Delta R.T. 0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21


Tgt Ion: 43 Resp: 27038

Ion Ratio Lower Upper

43 100

61 21.4 16.8 25.2



87 16.8 12.4 18.6




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0





#50

Toluene-d8

Concen: 52.308 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21

Instrument :

MSVOA_X

ClientSampleId :


PB116609ZHE#04

Tot Ion: 98 Resp: 1006188

Ion Ratio Lower Upper




98 100

100 64.3 52.0 78.0



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 52.473 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX007238.D

Acq: 25 Jan 2019 19:21


Tgt Ion: 95 Resp: 351890

Ion Ratio Lower Upper

95 100

174 94.8 0.0 182.2



176 92.6 0.0 178.8




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.01 min
Lab File: VX007238.D
Acq: 25 Jan 2019 19:21

Instrument :
MSVOA_X
ClientSampleId :
PB116609ZHE#04

Tot Ion:117 Resp: 800654
Ion Ratio Lower Upper
117 100
82 53.4 41.0 61.6
119 31.8 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX007238.D
Acq: 25 Jan 2019 19:21

Tgt Ion:152 Resp: 384985
Ion Ratio Lower Upper
152 100
115 54.3 39.1 117.2
150 157.1 0.0 344.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

