

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020746.D
 Acq On : 26 Jan 2021 12:23
 Operator : JC/SP
 Sample : M1192-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 26 14:39:50 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.623	168	282085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	474674	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	452025	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	197314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	187087	48.62	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.24%		
35) Dibromofluoromethane	5.464	113	157516	48.45	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.90%		
50) Toluene-d8	8.696	98	587272	49.27	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.54%		
62) 4-Bromofluorobenzene	11.122	95	210945	45.57	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.14%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.185	85	208	1.24	ug/l #	42
5) Bromomethane	1.624	94	1146	0.60	ug/l #	80
10) Methyl Iodide	2.483	142	2006	0.52	ug/l	92
11) Tert butyl alcohol	2.995	59	9838	13.70	ug/l #	76
13) Acrolein	2.288	56	275	0.51	ug/l #	47
15) Acrylonitrile	3.123	53	365	0.20	ug/l #	69
16) Acetone	2.428	43	6030	4.13	ug/l	93
17) Carbon Disulfide	2.538	76	2175	1.15	ug/l #	77
20) Methylene Chloride	2.819	84	788	0.21	ug/l #	64
25) 2-Butanone	4.635	43	729	0.32	ug/l #	88
29) Tetrahydrofuran	5.105	42	574	0.39	ug/l #	39
36) 1,1-Dichloropropene	5.623	75	21472	4.51	ug/l #	48
38) Carbon Tetrachloride	5.623	117	27927	5.58	ug/l #	17
52) Toluene	8.763	92	24031	2.72	ug/l	99
67) Ethyl Benzene	10.232	91	20008	1.18	ug/l	99
68) m/p-Xylenes	10.342	106	29683	4.69	ug/l	95
69) o-Xylene	10.683	106	16228	2.73	ug/l	96
73) Isopropylbenzene	11.000	105	2880	0.20	ug/l	98
74) N-amyl acetate	10.878	43	87	2.03	ug/l #	24
76) 1,2,3-Trichloropropane	11.122	75	104091	22.75	ug/l #	39
78) n-propylbenzene	11.341	91	10066	0.61	ug/l	95
79) 2-Chlorotoluene	11.415	91	6499	0.65	ug/l #	48
80) 1,3,5-Trimethylbenzene	11.488	105	16682	1.37	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.122	75	104091	63.74	ug/l #	12
82) 4-Chlorotoluene	11.488	91	2954	0.25	ug/l #	64
84) 1,2,4-Trimethylbenzene	11.793	105	54009	4.39	ug/l	99
86) p-Isopropyltoluene	12.006	119	2584	0.20	ug/l	97
89) n-Butylbenzene	12.366	91	6663	0.59	ug/l #	36
90) Hexachloroethane	12.573	117	1239	0.54	ug/l #	9
92) 1,2-Dibromo-3-Chloropr...	12.994	75	231	0.22	ug/l #	1
93) 1,2,4-Trichlorobenzene	13.628	180	905	0.22	ug/l #	51
94) Hexachlorobutadiene	13.762	225	591	0.31	ug/l	72
95) Naphthalene	13.817	128	18001	3.99	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 26 14:39:50 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.623 min Scan# 744
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 282085
Ion Ratio Lower Upper
168 100
99 60.5 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 1.24 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 85 Resp: 208
Ion Ratio Lower Upper
85 100
87 0.0 16.4 49.2#









#5
Bromomethane
Concen: 0.60 ug/l
RT: 1.624 min Scan# 88
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 1146
Ion Ratio Lower Upper
94 100
96 75.8 76.4 114.6#



#10
Methyl Iodide
Concen: 0.52 ug/l
RT: 2.483 min Scan# 229
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion:142 Resp: 2006
Ion Ratio Lower Upper
142 100
127 48.3 33.3 49.9
141 14.2 11.4 17.0









#11
Tert butyl alcohol
Concen: 13.70 ug/l
RT: 2.995 min Scan# 313
Delta R.T. -0.025 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 9838
Ion Ratio Lower Upper
59 100
57 0.9 7.8 11.8#



#13
Acrolein
Concen: 0.51 ug/l
RT: 2.288 min Scan# 197
Delta R.T. 0.018 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 56 Resp: 275
Ion Ratio Lower Upper
56 100
55 114.2 56.4 84.6#









#15
Acrylonitrile
Concen: 0.20 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 365
Ion Ratio Lower Upper
53 100
52 64.9 65.3 97.9#
51 0.0 28.2 42.2#



#16
Acetone
Concen: 4.13 ug/l
RT: 2.428 min Scan# 220
Delta R.T. 0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 43 Resp: 6030
Ion Ratio Lower Upper
43 100
58 39.1 27.8 41.8









#17
Carbon Disulfide
Concen: 1.15 ug/l
RT: 2.538 min Scan# 238
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 2175
Ion Ratio Lower Upper
76 100
78 17.4 7.2 10.8#



#20
Methylene Chloride
Concen: 0.21 ug/l
RT: 2.819 min Scan# 284
Delta R.T. -0.012 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 84 Resp: 788
Ion Ratio Lower Upper
84 100
49 57.0 88.6 132.8#
51 31.1 27.0 40.6
86 43.2 50.3 75.5#









#25
2-Butanone
Concen: 0.32 ug/l
RT: 4.635 min Scan# 582
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 729
Ion Ratio Lower Upper
43 100
72 35.9 23.7 35.5#



#29
Tetrahydrofuran
Concen: 0.39 ug/l
RT: 5.105 min Scan# 659
Delta R.T. 0.019 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 42 Resp: 574
Ion Ratio Lower Upper
42 100
72 3.7 41.7 62.5#
71 12.9 38.4 57.6#









#33
1,2-Dichloroethane-d4
Concen: 48.62 ug/l
RT: 6.025 min Scan# 810
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 65 Resp: 187087
Ion Ratio Lower Upper
65 100
67 52.7 0.0 105.4



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.824 min Scan# 941
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 114 Resp: 474674
Ion Ratio Lower Upper
114 100
63 21.6 0.0 40.8
88 17.2 0.0 32.8









#35
Dibromofluoromethane
Concen: 48.45 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:113 Resp: 157516
Ion Ratio Lower Upper
113 100
111 102.3 81.9 122.9
192 20.6 16.7 25.1



#36
1,1-Dichloropropene
Concen: 4.51 ug/l
RT: 5.623 min Scan# 744
Delta R.T. -0.146 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23


Tgt Ion: 75 Resp: 21472
Ion Ratio Lower Upper
75 100
110 8.7 18.1 54.4#
77 0.0 25.3 37.9#








#38
Carbon Tetrachloride
Concen: 5.58 ug/l
RT: 5.623 min Scan# 744
Delta R.T. -0.128 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion:117 Resp: 27927
Ion Ratio Lower Upper
117 100
119 6.0 76.2 114.2#
121 0.0 24.8 37.2#



#50
Toluene-d8
Concen: 49.27 ug/l
RT: 8.696 min Scan# 1248
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 98 Resp: 587272
Ion Ratio Lower Upper
98 100
100 65.0 51.9 77.9









#52
Toluene
Concen: 2.72 ug/l
RT: 8.763 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 92 Resp: 24031
Ion Ratio Lower Upper
92 100
91 169.7 137.1 205.7



#62
4-Bromofluorobenzene
Concen: 45.57 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 95 Resp: 210945
Ion Ratio Lower Upper
95 100
174 77.7 0.0 154.6
176 74.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.092 min Scan# 1477
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 452025
Ion Ratio Lower Upper
117 100
82 58.4 47.4 71.2
119 31.7 26.6 39.8



#67
Ethyl Benzene
Concen: 1.18 ug/l
RT: 10.232 min Scan# 1500
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 91 Resp: 20008
Ion Ratio Lower Upper
91 100
106 31.3 24.6 37.0









#68
m/p-Xylenes
Concen: 4.69 ug/l
RT: 10.342 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 29683
Ion Ratio Lower Upper
106 100
91 211.4 162.8 244.2



#69
o-Xylene
Concen: 2.73 ug/l
RT: 10.683 min Scan# 1574
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion:106 Resp: 16228
Ion Ratio Lower Upper
106 100
91 222.6 108.1 324.2









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:152 Resp: 197314
Ion Ratio Lower Upper
152 100
115 61.3 41.1 123.3
150 158.3 0.0 349.0



#73
Isopropylbenzene
Concen: 0.20 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23


Tgt Ion:105 Resp: 2880
Ion Ratio Lower Upper
105 100
120 25.3 13.2 39.6








#74
N-amy1 acetate
Concen: 2.03 ug/l
RT: 10.878 min Scan# 1606
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 43	Resp: 87
Ion Ratio	Lower Upper
43 100	
70 0.0	39.0 58.6#
55 81.6	21.5 32.3#
61 0.0	22.1 33.1#




#76
1,2,3-Trichloropropane
Concen: 22.75 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. -0.158 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23



Tgt Ion: 75	Resp: 104091
Ion Ratio	Lower Upper
75 100	
77 1.2	19.1 57.5#









#78
n-propylbenzene
Concen: 0.61 ug/l
RT: 11.341 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 10066
Ion Ratio Lower Upper
91 100
120 25.2 11.5 34.5



#79
2-Chlorotoluene
Concen: 0.65 ug/l
RT: 11.415 min Scan# 1694
Delta R.T. 0.013 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 91 Resp: 6499
Ion Ratio Lower Upper
91 100
126 3.5 16.6 49.8#









#80
1,3,5-Trimethylbenzene
Concen: 1.37 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 16682
Ion Ratio Lower Upper
105 100
120 48.5 23.8 71.5



#81
trans-1,4-Dichloro-2-butene
Concen: 63.74 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. 0.067 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 75 Resp: 104091
Ion Ratio Lower Upper
75 100
53 0.1 74.5 111.7#
89 0.0 37.9 56.9#









#82
4-Chlorotoluene
Concen: 0.25 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 2954
Ion Ratio Lower Upper
91 100
126 10.2 14.9 44.5#



#84
1,2,4-Trimethylbenzene
Concen: 4.39 ug/l
RT: 11.793 min Scan# 1756
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 105 Resp: 54009
Ion Ratio Lower Upper
105 100
120 44.3 22.4 67.3









#86
p-Isopropyltoluene
Concen: 0.20 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.043 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	119	Resp:	2584
Ion	Ratio	Lower	Upper
119	100		
134	26.4	12.9	38.7
91	25.9	11.7	35.0



#89
n-Butylbenzene
Concen: 0.59 ug/l
RT: 12.366 min Scan# 1850
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion:	91	Resp:	6663
Ion	Ratio	Lower	Upper
91	100		
92	34.4	27.4	82.0
134	96.8	13.2	39.5#









#90
Hexachloroethane
Concen: 0.54 ug/l
RT: 12.573 min Scan# 1884
Delta R.T. -0.006 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 1239
Ion Ratio Lower Upper
117 100
201 0.0 40.4 121.2#



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.22 ug/l
RT: 12.994 min Scan# 1953
Delta R.T. 0.013 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion: 75 Resp: 231
Ion Ratio Lower Upper
75 100
155 0.0 45.3 135.9#
157 0.0 57.7 173.1#









#93
1,2,4-Trichlorobenzene
Concen: 0.22 ug/l
RT: 13.628 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:180 Resp: 905
Ion Ratio Lower Upper
180 100
182 105.4 47.7 143.1
145 122.4 15.0 45.0#



#94
Hexachlorobutadiene
Concen: 0.31 ug/l
RT: 13.762 min Scan# 2079
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Tgt Ion:225 Resp: 591
Ion Ratio Lower Upper
225 100
223 42.3 31.1 93.2
227 39.8 31.6 95.0





#95
Naphthalene
Concen: 3.99 ug/l
RT: 13.817 min Scan# 2088
Delta R.T. -0.000 min
Lab File: VX020746.D
Acq: 26 Jan 2021 12:23

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 18001
Ion Ratio Lower Upper
128 100
127 14.1 10.3 15.5
129 12.0 8.8 13.2

