

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020755.D
 Acq On : 26 Jan 2021 15:54
 Operator : JC/SP
 Sample : M1189-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 41105

Quant Time: Jan 27 00:56:23 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	347799	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	599463	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	565098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	254489	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	228704	48.21	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.42%
35) Dibromofluoromethane	5.464	113	195673	47.65	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.30%
50) Toluene-d8	8.695	98	739449	49.12	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.24%
62) 4-Bromofluorobenzene	11.122	95	275802	47.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.36%

Target Compounds						Qvalue
11) Tert butyl alcohol	3.007	59	7958	8.99	ug/l	97
16) Acetone	2.416	43	31680	17.61	ug/l	92
18) Methyl Acetate	2.751	43	4246	0.92	ug/l	91
20) Methylene Chloride	2.837	84	1392	0.31	ug/l	91
25) 2-Butanone	4.641	43	6257	2.22	ug/l	94
40) Benzene	6.111	78	8055	0.45	ug/l	98
52) Toluene	8.762	92	9809	0.88	ug/l	96
68) m/p-Xylenes	10.335	106	1870	0.24	ug/l	92
84) 1,2,4-Trimethylbenzene	11.792	105	8170	0.52	ug/l	96
95) Naphthalene	13.816	128	3682	3.05	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
Data File : VX020755.D
Acq On : 26 Jan 2021 15:54
Operator : JC/SP
Sample : M1189-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
41105

Quant Time: Jan 27 00:56:23 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

