

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020763.D
 Acq On : 26 Jan 2021 19:01
 Operator : JC/SP
 Sample : M1217-19 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1674-1685

Quant Time: Jan 27 00:59:10 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	341792	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	583092	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	544648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	244276	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	219686	47.12	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.24%
35) Dibromofluoromethane	5.464	113	185751	46.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.02%
50) Toluene-d8	8.695	98	719807	49.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.32%
62) 4-Bromofluorobenzene	11.122	95	275375	48.43	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.86%
Target Compounds						
					Qvalue	
16) Acetone	2.422	43	3625	2.05	ug/l #	87
20) Methylene Chloride	2.831	84	1691	0.38	ug/l #	86
52) Toluene	8.762	92	2189	0.20	ug/l	99
91) 1,2-Dichlorobenzene	12.378	146	3230	0.41	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
Data File : VX020763.D
Acq On : 26 Jan 2021 19:01
Operator : JC/SP
Sample : M1217-19 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1674-1685

Quant Time: Jan 27 00:59:10 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

