

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020766.D
 Acq On : 26 Jan 2021 20:12
 Operator : JC/SP
 Sample : M1189-19 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 411087

Quant Time: Jan 27 01:00:19 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	347291	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	600184	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	594712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	287897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	225393	47.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.16%
35) Dibromofluoromethane	5.464	113	192793	46.90	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.80%
50) Toluene-d8	8.695	98	742272	49.25	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.50%
62) 4-Bromofluorobenzene	11.122	95	301803	51.56	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.12%
Target Compounds						
16) Acetone	2.416	43	2086	1.16	ug/l	94
84) 1,2,4-Trimethylbenzene	11.793	105	3849	0.21	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
Data File : VX020766.D
Acq On : 26 Jan 2021 20:12
Operator : JC/SP
Sample : M1189-19 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
411087

Quant Time: Jan 27 01:00:19 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

