

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020767.D
 Acq On : 26 Jan 2021 20:35
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 01:00:39 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	121	0.00
2 T	Dichlorodifluoromethane	50.000	44.946	10.1	108	0.00
3 P	Chloromethane	50.000	49.747	0.5	116	0.00
4 C	Vinyl Chloride	50.000	49.716	0.6#	114	0.00
5 T	Bromomethane	50.000	43.278	13.4	112	0.00
6 T	Chloroethane	50.000	51.021	-2.0	117	0.00
7 T	Trichlorofluoromethane	50.000	48.272	3.5	114	0.00
8 T	Diethyl Ether	50.000	51.425	-2.8	121	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.034	5.9	112	0.00
10 T	Methyl Iodide	50.000	50.169	-0.3	115	0.00
11 T	Tert butyl alcohol	250.000	253.489	-1.4	121	0.00
12 CM	1,1-Dichloroethene	50.000	49.046	1.9#	117	0.00
13 T	Acrolein	250.000	267.698	-7.1	125	0.00
14 T	Allyl chloride	50.000	50.122	-0.2	119	0.00
15 T	Acrylonitrile	250.000	257.697	-3.1	120	0.00
16 T	Acetone	250.000	242.988	2.8	118	0.00
17 T	Carbon Disulfide	50.000	45.633	8.7	113	0.00
18 T	Methyl Acetate	50.000	52.562	-5.1	120	0.00
19 T	Methyl tert-butyl Ether	50.000	53.560	-7.1	122	0.00
20 T	Methylene Chloride	50.000	48.521	3.0	120	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.622	2.8	117	0.00
22 T	Diisopropyl ether	50.000	52.347	-4.7	119	0.00
23 T	Vinyl Acetate	250.000	267.538	-7.0	120	0.00
24 P	1,1-Dichloroethane	50.000	50.477	-1.0	118	0.00
25 T	2-Butanone	250.000	259.081	-3.6	122	0.00
26 T	2,2-Dichloropropane	50.000	46.424	7.2	110	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.533	-3.1	121	0.00
28 T	Bromochloromethane	50.000	45.585	8.8	119	0.00
29 T	Tetrahydrofuran	250.000	264.481	-5.8	121	0.00
30 C	Chloroform	50.000	49.612	0.8#	116	0.00
31 T	Cyclohexane	50.000	50.373	-0.7	119	0.00
32 T	1,1,1-Trichloroethane	50.000	49.896	0.2	116	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.694	6.6	116	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	123	0.00
35 S	Dibromofluoromethane	50.000	47.019	6.0	118	0.00
36 T	1,1-Dichloropropene	50.000	47.795	4.4	118	0.00
37 T	Ethyl Acetate	50.000	51.927	-3.9	123	0.00
38 T	Carbon Tetrachloride	50.000	47.411	5.2	114	0.00
39 T	Methylcyclohexane	50.000	48.785	2.4	116	0.00
40 TM	Benzene	50.000	49.426	1.1	119	0.00
41 T	Methacrylonitrile	50.000	51.280	-2.6	122	0.00
42 TM	1,2-Dichloroethane	50.000	48.069	3.9	116	0.00
43 T	Isopropyl Acetate	50.000	52.421	-4.8	125	0.00
44 TM	Trichloroethene	50.000	48.710	2.6	119	0.00
45 C	1,2-Dichloropropane	50.000	50.369	-0.7#	121	0.00
46 T	Dibromomethane	50.000	49.184	1.6	118	0.00
47 T	Bromodichloromethane	50.000	48.235	3.5	116	0.00
48 T	Methyl methacrylate	50.000	52.879	-5.8	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020767.D
 Acq On : 26 Jan 2021 20:35
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 01:00:39 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	964.922	3.5	110	0.00
50 S	Toluene-d8	50.000	48.032	3.9	122	0.00
51 T	4-Methyl-2-Pentanone	250.000	263.186	-5.3	120	0.00
52 CM	Toluene	50.000	49.414	1.2#	119	0.00
53 T	t-1,3-Dichloropropene	50.000	49.888	0.2	120	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.726	-1.5	120	0.00
55 T	1,1,2-Trichloroethane	50.000	50.460	-0.9	123	0.00
56 T	Ethyl methacrylate	50.000	48.273	3.5	127	0.00
57 T	1,3-Dichloropropane	50.000	50.379	-0.8	121	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	263.050	-5.2	123	0.00
59 T	2-Hexanone	250.000	263.062	-5.2	118	0.00
60 T	Dibromochloromethane	50.000	48.802	2.4	118	0.00
61 T	1,2-Dibromoethane	50.000	50.421	-0.8	121	0.00
62 S	4-Bromofluorobenzene	50.000	47.309	5.4	117	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	124	0.00
64 T	Tetrachloroethene	50.000	50.527	-1.1	120	0.00
65 PM	Chlorobenzene	50.000	49.366	1.3	121	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.611	-1.2	121	0.00
67 C	Ethyl Benzene	50.000	51.355	-2.7#	120	0.00
68 T	m/p-Xylenes	100.000	102.680	-2.7	120	0.00
69 T	o-Xylene	50.000	52.827	-5.7	122	0.00
70 T	Styrene	50.000	52.729	-5.5	118	0.00
71 P	Bromoform	50.000	50.792	-1.6	119	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	115	0.00
73 T	Isopropylbenzene	50.000	54.809	-9.6	119	0.00
74 T	N-ethyl acetate	50.000	50.173	-0.3	119	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.078	-2.2	116	0.00
76 T	1,2,3-Trichloropropane	50.000	52.256	-4.5	115	0.00
77 T	Bromobenzene	50.000	52.268	-4.5	120	0.00
78 T	n-propylbenzene	50.000	53.301	-6.6	116	0.00
79 T	2-Chlorotoluene	50.000	52.240	-4.5	115	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.150	-8.3	116	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.691	-3.4	117	0.00
82 T	4-Chlorotoluene	50.000	51.008	-2.0	113	0.00
83 T	tert-Butylbenzene	50.000	53.420	-6.8	115	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.378	-8.8	116	0.00
85 T	sec-Butylbenzene	50.000	54.334	-8.7	116	0.00
86 T	p-Isopropyltoluene	50.000	53.675	-7.3	115	0.00
87 T	1,3-Dichlorobenzene	50.000	50.078	-0.2	112	0.00
88 T	1,4-Dichlorobenzene	50.000	47.987	4.0	113	0.00
89 T	n-Butylbenzene	50.000	51.551	-3.1	110	0.00
90 T	Hexachloroethane	50.000	49.735	0.5	109	0.00
91 T	1,2-Dichlorobenzene	50.000	49.585	0.8	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.108	-0.2	118	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.239	-4.5	117	0.00
94 T	Hexachlorobutadiene	50.000	44.071	11.9	103	0.00
95 T	Naphthalene	50.000	49.410	1.2	120	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
Data File : VX020767.D
Acq On : 26 Jan 2021 20:35
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 27 01:00:39 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.752	-9.5	122	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6