



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012621\
 Data File : VX020740.D
 Acq On : 26 Jan 2021 08:45
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Title : SW846 8260
 Last Update : Mon Jan 25 13:38:15 2021



AutoFind: Scans 1644, 1645, 1646; Background Corrected with Scan 1637

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	4040	PASS
75	95	30	60	47.3	11434	PASS
95	95	100	100	100.0	24155	PASS
96	95	5	9	6.7	1626	PASS
173	174	0.00	2	0.9	157	PASS
174	95	50	100	72.0	17392	PASS
175	174	5	9	7.8	1351	PASS
176	174	95	101	99.4	17283	PASS
177	176	5	9	6.1	1060	PASS