

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012623\
 Data File : VX033904.D
 Acq On : 26 Jan 2023 12:28
 Operator : JC/MD
 Sample : VX0126WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0126WBL01

Quant Time: Jan 27 02:52:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	225419	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90108	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74388	45.468	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.940%
35) Dibromofluoromethane	5.385	113	69173	49.419	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.840%
50) Toluene-d8	8.647	98	251860	49.385	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.760%
62) 4-Bromofluorobenzene	11.079	95	84650	44.705	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012623\
Data File : VX033904.D
Acq On : 26 Jan 2023 12:28
Operator : JC/MD
Sample : VX0126WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0126WBL01

Quant Time: Jan 27 02:52:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

