

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012623\
 Data File : VX033909.D
 Acq On : 26 Jan 2023 15:09
 Operator : JC/MD
 Sample : 01323-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jan 27 02:55:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	236278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99218	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79142	45.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.940%
35) Dibromofluoromethane	5.385	113	73374	50.011	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.020%
50) Toluene-d8	8.647	98	270543	50.610	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.220%
62) 4-Bromofluorobenzene	11.079	95	91938	46.323	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012623\
Data File : VX033909.D
Acq On : 26 Jan 2023 15:09
Operator : JC/MD
Sample : 01323-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jan 27 02:55:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

