

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012623\
 Data File : VX033918.D
 Acq On : 26 Jan 2023 18:38
 Operator : JC/MD
 Sample : 01327-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N48965

Quant Time: Jan 27 02:59:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	150244	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	218704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73960	46.753	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.500%
35) Dibromofluoromethane	5.379	113	68035	50.098	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.200%
50) Toluene-d8	8.647	98	249742	50.473	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.940%
62) 4-Bromofluorobenzene	11.079	95	87654	47.713	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.420%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	636	0.500	ug/l	89
27) cis-1,2-Dichloroethene	4.495	96	2257	1.374	ug/l	90
40) Benzene	6.043	78	2314	0.401	ug/l	97
44) Trichloroethene	7.129	130	760	0.470	ug/l	95
52) Toluene	8.720	92	1865	0.500	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012623\
Data File : VX033918.D
Acq On : 26 Jan 2023 18:38
Operator : JC/MD
Sample : 01327-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N48965

Quant Time: Jan 27 02:59:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

