

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
 Data File : VX039976.D
 Acq On : 26 Jan 2024 12:29
 Operator : JC/MD
 Sample : P1263-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-012524

Quant Time: Jan 29 00:39:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50914	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50402	44.635	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.280%
35) Dibromofluoromethane	5.385	113	39655	48.631	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.647	98	151901	49.391	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.780%
62) 4-Bromofluorobenzene	11.079	95	58432	48.586	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.180%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
Data File : VX039976.D
Acq On : 26 Jan 2024 12:29
Operator : JC/MD
Sample : P1263-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-012524

Quant Time: Jan 29 00:39:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

