

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
 Data File : VX039978.D
 Acq On : 26 Jan 2024 13:15
 Operator : JC/MD
 Sample : P1263-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-08B-72.5-012524

Quant Time: Jan 29 00:40:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	42280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	83248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	78963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38928	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	39896	56.330	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.660%	
35) Dibromofluoromethane	5.391	113	28244	50.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.740%	
50) Toluene-d8	8.653	98	107178	50.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.360%	
62) 4-Bromofluorobenzene	11.079	95	41293	49.933	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
Data File : VX039978.D
Acq On : 26 Jan 2024 13:15
Operator : JC/MD
Sample : P1263-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-08B-72.5-012524

Quant Time: Jan 29 00:40:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

