

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
 Data File : VX039979.D
 Acq On : 26 Jan 2024 14:13
 Operator : JC/MD
 Sample : P1261-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-458-460

Quant Time: Jan 29 00:40:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	54358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	98393	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	84593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	41781	45.884	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.760%
35) Dibromofluoromethane	5.385	113	31447	47.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.900%
50) Toluene-d8	8.647	98	117010	46.814	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.620%
62) 4-Bromofluorobenzene	11.079	95	45362	46.410	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
16) Acetone	2.380	43	1228	2.398	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
Data File : VX039979.D
Acq On : 26 Jan 2024 14:13
Operator : JC/MD
Sample : P1261-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-458-460

Quant Time: Jan 29 00:40:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

