

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
 Data File : VX039981.D
 Acq On : 26 Jan 2024 15:01
 Operator : JC/MD
 Sample : P1261-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-EB-20240124

Quant Time: Jan 29 00:41:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	43651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	87902	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	83304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	38504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	40232	55.021	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	110.040%		
35) Dibromofluoromethane	5.391	113	28512	48.157	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.320%		
50) Toluene-d8	8.647	98	114197	51.141	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	102.280%		
62) 4-Bromofluorobenzene	11.079	95	39856	45.643	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.280%		
Target Compounds						
16) Acetone	2.380	43	932	2.266	ug/l #	78
29) Tetrahydrofuran	5.013	42	4966	13.056	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
Data File : VX039981.D
Acq On : 26 Jan 2024 15:01
Operator : JC/MD
Sample : P1261-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-EB-20240124

Quant Time: Jan 29 00:41:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

