

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
 Data File : VX039983.D
 Acq On : 26 Jan 2024 15:47
 Operator : JC/MD
 Sample : P1261-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-498-500

Quant Time: Jan 29 00:42:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	45950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	87994	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	86082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43537	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	41025	53.298	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.600%
35) Dibromofluoromethane	5.391	113	29589	49.924	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.840%
50) Toluene-d8	8.647	98	117961	52.771	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.540%
62) 4-Bromofluorobenzene	11.079	95	44544	50.959	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
Data File : VX039983.D
Acq On : 26 Jan 2024 15:47
Operator : JC/MD
Sample : P1261-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-498-500

Quant Time: Jan 29 00:42:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

