

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
 Data File : VX039988.D
 Acq On : 26 Jan 2024 17:42
 Operator : JC/MD
 Sample : P1262-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-505-507

Quant Time: Jan 29 00:44:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	38956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	74163	50.000	ug/l	0.01
63) Chlorobenzene-d5	10.055	117	70967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	32141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	39226	60.110	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 120.220%#	
35) Dibromofluoromethane	5.385	113	24899	49.846	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.700%	
50) Toluene-d8	8.647	98	94287	50.047	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	11.079	95	35777	48.562	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.120%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012624\
Data File : VX039988.D
Acq On : 26 Jan 2024 17:42
Operator : JC/MD
Sample : P1262-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-505-507

Quant Time: Jan 29 00:44:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

