

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012720\
 Data File : VX014720.D
 Acq On : 27 Jan 2020 12:48
 Operator : JC/SP
 Sample : L1264-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CTY-GW-01-12

Quant Time: Jan 28 04:40:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	450810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	703626	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	644174	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	293987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	258512	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
35) Dibromofluoromethane	5.49	113	217269	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
50) Toluene-d8	8.71	98	844857	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
62) 4-Bromofluorobenzene	11.13	95	283040	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
Target Compounds						
					Qvalue	
16) Acetone	2.43	43	22182	7.834	ug/l	95
39) Methylcyclohexane	7.46	83	3236	0.395	ug/l	90
40) Benzene	6.14	78	58584	2.900	ug/l	98
67) Ethyl Benzene	10.25	91	7415	0.315	ug/l	98
68) m/p-Xylenes	10.35	106	4750	0.533	ug/l	78
69) o-Xylene	10.70	106	3381	0.393	ug/l	88
84) 1,2,4-Trimethylbenzene	11.81	105	6620	0.385	ug/l	95
95) Naphthalene	13.83	128	44505	3.051	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012720\
Data File : VX014720.D
Acq On : 27 Jan 2020 12:48
Operator : JC/SP
Sample : L1264-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CTY-GW-01-12

Quant Time: Jan 28 04:40:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

