

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012720\
 Data File : VX014732.D
 Acq On : 27 Jan 2020 17:27
 Operator : JC/SP
 Sample : L1245-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20200067-AMENDED-COMP-5

Quant Time: Jan 28 05:55:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	453424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	718954	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	663756	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	309138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	263409	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
35) Dibromofluoromethane	5.49	113	119373	26.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.62%	
50) Toluene-d8	8.71	98	861658	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
62) 4-Bromofluorobenzene	11.13	95	296577	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
Target Compounds						
16) Acetone	2.43	43	57452	19.936	ug/l	98
20) Methylene Chloride	2.85	84	12498	2.327	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012720\
Data File : VX014732.D
Acq On : 27 Jan 2020 17:27
Operator : JC/SP
Sample : L1245-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20200067-AMENDED-COMP-5

Quant Time: Jan 28 05:55:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

