


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX012720\
 Data File : VX014713.D
 Acq On : 27 Jan 2020 09:15
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Title : SW846 8260
 Last Update : Wed Jan 22 12:35:15 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	18201	PASS
75	95	30	60	51.0	43383	PASS
95	95	100	100	100.0	85138	PASS
96	95	5	9	6.8	5752	PASS
173	174	0.00	2	1.4	999	PASS
174	95	50	100	85.2	72549	PASS
175	174	5	9	7.3	5303	PASS
176	174	95	101	96.0	69674	PASS
177	176	5	9	7.1	4920	PASS