

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012721\
 Data File : VX020773.D
 Acq On : 27 Jan 2021 16:27
 Operator : JC/SP
 Sample : M1189-20 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 41088

Quant Time: Jan 28 00:44:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	313820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	525548	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	506656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	224827	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	204750	47.83	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.66%
35) Dibromofluoromethane	5.464	113	175211	48.67	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.34%
50) Toluene-d8	8.695	98	644045	48.80	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.60%
62) 4-Bromofluorobenzene	11.122	95	236270	46.10	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.20%
Target Compounds						
95) Naphthalene	13.823	128	1669	2.96	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012721\
Data File : VX020773.D
Acq On : 27 Jan 2021 16:27
Operator : JC/SP
Sample : M1189-20 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
41088

Quant Time: Jan 28 00:44:47 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

