

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012721\
 Data File : VX020783.D
 Acq On : 27 Jan 2021 20:14
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 28 00:48:08 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	411704	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	709356	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	670543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	314650	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.032	65	269100	47.92	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.84%
35) Dibromofluoromethane	5.458	113	231468	47.64	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.28%
50) Toluene-d8	8.696	98	869396	48.81	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.62%
62) 4-Bromofluorobenzene	11.116	95	330093	47.72	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.44%
Target Compounds						
20) Methylene Chloride	2.837	84	1507	0.28	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012721\
Data File : VX020783.D
Acq On : 27 Jan 2021 20:14
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 28 00:48:08 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

