

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026587.D
 Acq On : 26 Jan 2022 12:24
 Operator : JC/MD
 Sample : VX0127WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBSD01

Quant Time: Jan 27 07:20:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	155743	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65653	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63122	46.705	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.420%
35) Dibromofluoromethane	5.391	113	49858	49.573	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.140%
50) Toluene-d8	8.647	98	185865	51.284	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.560%
62) 4-Bromofluorobenzene	11.079	95	70373	54.660	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.320%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.172	85	18977	16.611	ug/l	98
3) Chloromethane	1.294	50	19312	18.440	ug/l	99
4) Vinyl Chloride	1.380	62	19865	18.066	ug/l	99
5) Bromomethane	1.617	94	11228	22.136	ug/l	94
6) Chloroethane	1.691	64	10571	19.012	ug/l	100
7) Trichlorofluoromethane	1.892	101	30027	17.297	ug/l	99
8) Diethyl Ether	2.136	74	11637	18.943	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.337	101	19937	18.845	ug/l	100
10) Methyl Iodide	2.459	142	25338	24.715	ug/l	96
11) Tert butyl alcohol	3.001	59	15407	81.949	ug/l	99
12) 1,1-Dichloroethene	2.325	96	18579	18.388	ug/l	91
13) Acrolein	2.245	56	19032	95.266	ug/l	99
14) Allyl chloride	2.666	41	34645	18.645	ug/l	96
15) Acrylonitrile	3.068	53	59688	103.134	ug/l	99
16) Acetone	2.392	43	64010	92.874	ug/l	93
17) Carbon Disulfide	2.514	76	47879	16.230	ug/l	99
18) Methyl Acetate	2.709	43	28039	20.464	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	67841	18.504	ug/l	98
20) Methylene Chloride	2.794	84	22012	19.330	ug/l	97
21) trans-1,2-Dichloroethene	3.099	96	20255	18.517	ug/l	95
22) Diisopropyl ether	3.757	45	72311	19.158	ug/l #	84
23) Vinyl Acetate	3.727	43	298092	95.945	ug/l	98
24) 1,1-Dichloroethane	3.611	63	37979	18.632	ug/l	99
25) 2-Butanone	4.568	43	86595	96.614	ug/l	99
26) 2,2-Dichloropropane	4.483	77	26963	17.845	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	24152	19.347	ug/l	98
28) Bromochloromethane	4.897	49	15789	18.864	ug/l	97
29) Tetrahydrofuran	5.013	42	53360	100.239	ug/l	100
30) Chloroform	5.099	83	39961	18.868	ug/l	99
31) Cyclohexane	5.477	56	34968	17.542	ug/l	99
32) 1,1,1-Trichloroethane	5.391	97	33938	18.217	ug/l	99
36) 1,1-Dichloropropene	5.690	75	29179	18.378	ug/l	98
37) Ethyl Acetate	4.721	43	32055	19.959	ug/l	99
38) Carbon Tetrachloride	5.678	117	30322	18.567	ug/l	97
39) Methylcyclohexane	7.385	83	36091	18.846	ug/l	98
40) Benzene	6.044	78	85062	19.383	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026587.D
 Acq On : 26 Jan 2022 12:24
 Operator : JC/MD
 Sample : VX0127WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBSD01

Quant Time: Jan 27 07:20:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	18345	20.164	ug/l	98
42) 1,2-Dichloroethane	6.092	62	32922	19.174	ug/l	99
43) Isopropyl Acetate	6.342	43	51607	19.892	ug/l	99
44) Trichloroethene	7.129	130	24674	19.815	ug/l	98
45) 1,2-Dichloropropane	7.434	63	21719	19.827	ug/l	100
46) Dibromomethane	7.580	93	15628	20.239	ug/l	99
47) Bromodichloromethane	7.824	83	31242	20.024	ug/l	100
48) Methyl methacrylate	7.696	41	25337	19.706	ug/l	97
49) 1,4-Dioxane	7.696	88	7391	350.291	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	164893	107.156	ug/l	99
52) Toluene	8.720	92	55380	20.579	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	30510	19.484	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	34549	20.617	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	23019	21.919	ug/l	99
56) Ethyl methacrylate	9.116	69	34393	21.370	ug/l	99
57) 1,3-Dichloropropane	9.311	76	38740	21.277	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	89469	106.017	ug/l	100
59) 2-Hexanone	9.433	43	122824	104.285	ug/l	98
60) Dibromochloromethane	9.519	129	23792	21.741	ug/l	99
61) 1,2-Dibromoethane	9.610	107	23120	21.128	ug/l	98
64) Tetrachloroethene	9.275	164	25839	19.351	ug/l	94
65) Chlorobenzene	10.079	112	59257	19.907	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	21888	19.790	ug/l	99
67) Ethyl Benzene	10.195	91	106020	19.600	ug/l	99
68) m/p-Xylenes	10.305	106	81019	40.023	ug/l	100
69) o-Xylene	10.640	106	39615	20.281	ug/l	99
70) Styrene	10.659	104	66936	21.000	ug/l	98
71) Bromoform	10.799	173	15913	20.200	ug/l #	100
73) Isopropylbenzene	10.963	105	106781	19.842	ug/l	99
74) N-amyl acetate	10.841	43	40517	19.502	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	28860	19.980	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	33735	23.337	ug/l	92
77) Bromobenzene	11.201	156	25128	20.382	ug/l	98
78) n-propylbenzene	11.305	91	121509	19.800	ug/l	100
79) 2-Chlorotoluene	11.366	91	73728	19.812	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	88949	20.049	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	7910	20.518	ug/l	99
82) 4-Chlorotoluene	11.457	91	84207	19.688	ug/l	99
83) tert-Butylbenzene	11.713	119	84163	20.321	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	88713	19.992	ug/l	100
85) sec-Butylbenzene	11.890	105	108557	20.264	ug/l	99
86) p-Isopropyltoluene	12.012	119	91269	20.307	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	46044	19.925	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	46142	19.727	ug/l	98
89) n-Butylbenzene	12.335	91	76947	19.950	ug/l	99
90) Hexachloroethane	12.542	117	13181	18.986	ug/l	95
91) 1,2-Dichlorobenzene	12.335	146	43819	20.065	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6329	18.129	ug/l	95
93) 1,2,4-Trichlorobenzene	13.591	180	27037	19.311	ug/l	98
94) Hexachlorobutadiene	13.725	225	11341	19.189	ug/l	99
95) Naphthalene	13.780	128	90507	19.380	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	27211	19.762	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026587.D
Acq On : 26 Jan 2022 12:24
Operator : JC/MD
Sample : VX0127WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Quant Time: Jan 27 07:20:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jan 27 07:20:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01







Tgt Ion:168 Resp: 96748
Ion Ratio Lower Upper
168 100
99 58.1 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 16.611 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 85 Resp: 18977
Ion Ratio Lower Upper
85 100
87 31.7 16.4 49.1









#3
Chloromethane
Concen: 18.440 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 50 Resp: 19312
Ion Ratio Lower Upper
50 100
52 31.9 26.1 39.1



#4
Vinyl Chloride
Concen: 18.066 ug/l
RT: 1.380 min Scan# 48
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 62 Resp: 19865
Ion Ratio Lower Upper
62 100
64 32.1 26.2 39.2









#5
Bromomethane
Concen: 22.136 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 94 Resp: 11228
Ion Ratio Lower Upper
94 100
96 100.3 75.9 113.9



#6
Chloroethane
Concen: 19.012 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 64 Resp: 10571
Ion Ratio Lower Upper
64 100
66 32.0 25.4 38.2









#7
 Trichlorofluoromethane
 Concen: 17.297 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBSD01


Tgt Ion:101 Resp: 30027
 Ion Ratio Lower Upper
 101 100
 103 64.8 51.3 76.9



#8
 Diethyl Ether
 Concen: 18.943 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24

Tgt Ion: 74 Resp: 11637
 Ion Ratio Lower Upper
 74 100
 45 96.8 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.845 ug/l
RT: 2.337 min Scan# 205
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01

Tgt Ion:101 Resp: 19937




Ion	Ratio	Lower	Upper
101	100		
85	44.2	35.3	52.9
151	76.6	61.4	92.2




#10
Methyl Iodide
Concen: 24.715 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:142 Resp: 25338

Ion	Ratio	Lower	Upper
142	100		
127	44.8	38.1	57.1
141	14.5	11.2	16.8









#11
Tert butyl alcohol
Concen: 81.949 ug/l
RT: 3.001 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01

Tgt Ion: 59 Resp: 15407
Ion Ratio Lower Upper
59 100
57 9.3 7.2 10.8




#12
1,1-Dichloroethene
Concen: 18.388 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 96 Resp: 18579
Ion Ratio Lower Upper
96 100
61 158.2 138.5 207.7
98 61.5 51.8 77.6




Abundance Scan 189 (2.239 min): VX026417.D\data.ms (-18




m/z-->

Abundance Scan 190 (2.245 min): VX026587.D\data.ms



m/z-->

Abundance Scan 190 (2.245 min): VX026587.D\data.ms (-14



m/z-->

#13

Acrolein

Concen: 95.266 ug/l

RT: 2.245 min Scan# 190

Delta R.T. 0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01


Tgt Ion: 56 Resp: 19032

Ion Ratio Lower Upper


56 100

55 70.6 56.9 85.3

Abundance




Abundance Scan 260 (2.672 min): VX026417.D\data.ms (-25




m/z-->

Abundance Scan 259 (2.666 min): VX026587.D\data.ms



m/z-->

Abundance Scan 259 (2.666 min): VX026587.D\data.ms (-21



m/z-->

#14

Allyl chloride

Concen: 18.645 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Tgt Ion: 41 Resp: 34645


Ion Ratio Lower Upper


41 100

39 63.8 54.2 81.4

76 31.2 25.2 37.8

Abundance









#15
Acrylonitrile
Concen: 103.134 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 53	Resp:	59688
Ion	Ratio	Lower Upper
53	100	
52	82.9	66.6 100.0
51	36.5	29.5 44.3



#16
Acetone
Concen: 92.874 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 43	Resp:	64010
Ion	Ratio	Lower Upper
43	100	
58	32.3	23.0 34.6









#17
Carbon Disulfide
Concen: 16.230 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 76 Resp: 47879
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3



#18
Methyl Acetate
Concen: 20.464 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 43 Resp: 28039
Ion Ratio Lower Upper
43 100
74 22.8 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 18.504 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 73 Resp: 67841
Ion Ratio Lower Upper
73 100
57 22.8 17.4 26.2



#20
Methylene Chloride
Concen: 19.330 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 84 Resp: 22012
Ion Ratio Lower Upper
84 100
49 122.5 101.5 152.3
51 37.8 31.2 46.8
86 65.0 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 18.517 ug/l
RT: 3.099 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 96 Resp: 20255
Ion Ratio Lower Upper
96 100
61 132.9 112.9 169.3
98 63.9 49.5 74.3



#22
Diisopropyl ether
Concen: 19.158 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 45 Resp: 72311
Ion Ratio Lower Upper
45 100
43 96.1 60.6 91.0#
87 26.4 22.0 33.0
59 11.3 9.0 13.4









#23
Vinyl Acetate
Concen: 95.945 ug/l
RT: 3.727 min Scan# 41
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 43 Resp: 298092
Ion Ratio Lower Upper
43 100
86 10.7 8.1 12.1



#24
1,1-Dichloroethane
Concen: 18.632 ug/l
RT: 3.611 min Scan# 414
Delta R.T. -0.006 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 63 Resp: 37979
Ion Ratio Lower Upper
63 100
98 8.1 3.7 11.1
100 4.8 2.4 7.2





#25

2-Butanone

Concen: 96.614 ug/l

RT: 4.568 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 43 Resp: 86595

Ion Ratio Lower Upper

43 100

72 25.1 20.4 30.6



#26

2,2-Dichloropropane

Concen: 17.845 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 77 Resp: 26963

Ion Ratio Lower Upper

77 100

97 23.2 11.0 33.0









#27
 cis-1,2-Dichloroethene
 Concen: 19.347 ug/l
 RT: 4.495 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBSD01


Tgt Ion: 96 Resp: 24152
 Ion Ratio Lower Upper
 96 100
 61 145.7 0.0 296.6
 98 65.2 0.0 128.8



#28
 Bromochloromethane
 Concen: 18.864 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24

Tgt Ion: 49 Resp: 15789
 Ion Ratio Lower Upper
 49 100
 129 1.2 0.0 4.2
 130 80.5 62.4 93.6









#29
Tetrahydrofuran
Concen: 100.239 ug/l
RT: 5.013 min Scan# 644
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:	42	Resp:	53360
Ion	Ratio	Lower	Upper
42	100		
72	43.4	34.6	52.0
71	40.0	31.8	47.8



#30
Chloroform
Concen: 18.868 ug/l
RT: 5.099 min Scan# 658
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:	83	Resp:	39961
Ion	Ratio	Lower	Upper
83	100		
85	65.5	51.6	77.4









#31
Cyclohexane
Concen: 17.542 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:	56	Resp:	34968
Ion	Ratio	Lower	Upper
56	100		
69	31.2	25.8	38.6
84	86.8	68.9	103.3



#32
1,1,1-Trichloroethane
Concen: 18.217 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:	97	Resp:	33938
Ion	Ratio	Lower	Upper
97	100		
99	64.7	51.3	76.9
61	44.5	35.0	52.6





#33

1,2-Dichloroethane-d4

Concen: 46.705 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 65 Resp: 63122

Ion Ratio Lower Upper

65 100

67 51.3 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24




Tgt Ion: 114 Resp: 155743


Ion Ratio Lower Upper

114 100

63 21.5 0.0 42.8

88 16.1 0.0 33.0





#35

Dibromofluoromethane

Concen: 49.573 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01





Tgt Ion:113 Resp: 49858

Ion Ratio Lower Upper

113 100

111 102.5 81.2 121.8

192 19.0 15.0 22.4



#36

1,1-Dichloropropene

Concen: 18.378 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24




Tgt Ion: 75 Resp: 29179


Ion Ratio Lower Upper

75 100

110 35.6 17.2 51.6

77 31.7 24.6 36.8





#37

Ethyl Acetate

Concen: 19.959 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01


Tgt Ion: 43 Resp: 32055

Ion Ratio Lower Upper




43 100

61 13.6 10.6 16.0

70 10.7 8.5 12.7



Abundance



#38

Carbon Tetrachloride

Concen: 18.567 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24


Tgt Ion: 117 Resp: 30322

Ion Ratio Lower Upper



117 100


119 96.0 74.7 112.1

121 29.6 25.4 38.0



Abundance





#39

Methylcyclohexane

Concen: 18.846 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01





Tgt Ion: 83 Resp: 36091

Ion Ratio Lower Upper

83 100

55 77.8 63.6 95.4

98 50.3 38.4 57.6



#40

Benzene

Concen: 19.383 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 78 Resp: 85062

Ion Ratio Lower Upper

78 100

77 22.5 18.9 28.3









#41
Methacrylonitrile
Concen: 20.164 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01

Tgt Ion: 41 Resp: 18345




Ion	Ratio	Lower	Upper
41	100		
39	54.8	45.4	68.0
67	68.7	53.2	79.8
52	28.7	23.0	34.6




#42
1,2-Dichloroethane
Concen: 19.174 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 62 Resp: 32922

Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	17.8





#43

Isopropyl Acetate

Concen: 19.892 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01





Tgt Ion: 43 Resp: 51607

Ion Ratio Lower Upper

43 100

61 19.3 15.4 23.0

87 13.2 10.2 15.2



#44

Trichloroethene

Concen: 19.815 ug/l

RT: 7.129 min Scan# 991

Delta R.T. 0.000 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 130 Resp: 24674

Ion Ratio Lower Upper

130 100

95 98.9 0.0 202.6





#45

1,2-Dichloropropane

Concen: 19.827 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 63 Resp: 21719

Ion Ratio Lower Upper

63 100

65 29.8 23.8 35.6



#46

Dibromomethane

Concen: 20.239 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24




Tgt Ion: 93 Resp: 15628


Ion Ratio Lower Upper

93 100

95 83.8 66.1 99.1

174 99.3 78.7 118.1









#47
Bromodichloromethane
Concen: 20.024 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 83	Resp:	31242	
Ion Ratio	Lower	Upper	
83	100		
85	64.1	51.4	77.0
127	9.1	6.9	10.3



#48
Methyl methacrylate
Concen: 19.706 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 41	Resp: 25337
Ion Ratio	Lower Upper
41 100	
69 80.0	62.2 93.4
39 54.7	45.4 68.0





#49

1,4-Dioxane

Concen: 350.291 ug/l

RT: 7.696 min Scan# 110

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01





Tgt Ion: 88 Resp: 7391

Ion Ratio Lower Upper

88 100

43 35.7 29.9 44.9

58 72.2 57.8 86.6



#50

Toluene-d8

Concen: 51.284 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 98 Resp: 185865

Ion Ratio Lower Upper

98 100

100 65.9 53.2 79.8





#51

4-Methyl-2-Pentanone

Concen: 107.156 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 43 Resp: 164893

Ion Ratio Lower Upper

43 100

58 38.5 30.5 45.7



#52

Toluene

Concen: 20.579 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 92 Resp: 55380

Ion Ratio Lower Upper

92 100

91 169.9 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 19.484 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 75 Resp: 30510

Ion Ratio Lower Upper

75 100

77 30.5 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 20.617 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24




Tgt Ion: 75 Resp: 34549


Ion Ratio Lower Upper

75 100

77 33.4 26.2 39.2

39 51.3 43.5 65.3









#55
1,1,2-Trichloroethane
Concen: 21.919 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01

Tgt Ion: 97 Resp: 23019




Ion	Ratio	Lower	Upper
97	100		
83	85.4	68.4	102.6
85	56.2	44.0	66.0
99	62.4	49.8	74.6




#56
Ethyl methacrylate
Concen: 21.370 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 69 Resp: 34393

Ion	Ratio	Lower	Upper
69	100		
41	70.7	57.2	85.8
39	38.3	30.9	46.3









#57
1,3-Dichloropropane
Concen: 21.277 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 76 Resp: 38740
Ion Ratio Lower Upper
76 100
78 32.1 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 106.017 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 63 Resp: 89469
Ion Ratio Lower Upper
63 100
106 29.0 23.0 34.6





#59

2-Hexanone

Concen: 104.285 ug/l

RT: 9.433 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 43 Resp: 122824

Ion Ratio Lower Upper

43 100

58 53.9 26.4 79.2



#60

Dibromochloromethane

Concen: 21.741 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.006 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 129 Resp: 23792

Ion Ratio Lower Upper

129 100

127 78.0 38.5 115.5









#61
1,2-Dibromoethane
Concen: 21.128 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:107 Resp: 23120
Ion Ratio Lower Upper
107 100
109 96.1 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 54.660 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 95 Resp: 70373
Ion Ratio Lower Upper
95 100
174 74.0 0.0 148.6
176 73.0 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:117 Resp: 141352
Ion Ratio Lower Upper
117 100
82 55.9 45.2 67.8
119 32.5 26.2 39.2



#64
Tetrachloroethene
Concen: 19.351 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:164 Resp: 25839
Ion Ratio Lower Upper
164 100
166 133.0 101.8 152.6
129 94.0 83.3 124.9
131 93.4 78.0 117.0









#65
Chlorobenzene
Concen: 19.907 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:112 Resp: 59257
Ion Ratio Lower Upper
112 100
114 31.7 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 19.790 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:131 Resp: 21888
Ion Ratio Lower Upper
131 100
133 95.2 47.9 143.8
119 64.6 31.7 95.1





#67

Ethyl Benzene

Concen: 19.600 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :





VX0127WBSD01

Tgt Ion: 91 Resp: 106020

Ion Ratio Lower Upper

91 100

106 31.3 24.7 37.1



#68

m/p-Xylenes

Concen: 40.023 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion:106 Resp: 81019

Ion Ratio Lower Upper

106 100

91 206.6 165.3 247.9









#69
o-Xylene
Concen: 20.281 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:106 Resp: 39615
Ion Ratio Lower Upper
106 100
91 218.0 109.7 329.0



#70
Styrene
Concen: 21.000 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:104 Resp: 66936
Ion Ratio Lower Upper
104 100
78 50.5 42.5 63.7
103 54.5 43.3 64.9









#71
Bromoform
Concen: 20.200 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:173 Resp: 15913
Ion Ratio Lower Upper
173 100
175 48.9 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:152 Resp: 65653
Ion Ratio Lower Upper
152 100
115 70.3 44.2 132.6
150 164.1 0.0 348.2









#73
Isopropylbenzene
Concen: 19.842 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:105 Resp: 106781
Ion Ratio Lower Upper
105 100
120 25.7 13.1 39.1



#74
N-ethyl acetate
Concen: 19.502 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 43 Resp: 40517
Ion Ratio Lower Upper
43 100
70 43.7 32.9 49.3
55 25.5 18.3 27.5
61 23.6 18.9 28.3









#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.980 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.000 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBSD01


Tgt Ion: 83 Resp: 28860
 Ion Ratio Lower Upper
 83 100
 131 9.8 5.3 15.9
 85 65.3 32.0 96.2



#76
 1,2,3-Trichloropropane
 Concen: 23.337 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24

Tgt Ion: 75 Resp: 33735
 Ion Ratio Lower Upper
 75 100
 77 32.0 18.3 54.8









#77
Bromobenzene
Concen: 20.382 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:156 Resp: 25128
Ion Ratio Lower Upper
156 100
77 164.0 84.5 253.4
158 96.0 48.1 144.4



#78
n-propylbenzene
Concen: 19.800 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 91 Resp: 121509
Ion Ratio Lower Upper
91 100
120 22.9 11.6 34.7









#79
2-Chlorotoluene
Concen: 19.812 ug/l
RT: 11.366 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 91 Resp: 73728
Ion Ratio Lower Upper
91 100
126 33.7 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 20.049 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 105 Resp: 88949
Ion Ratio Lower Upper
105 100
120 48.1 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 20.518 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion: 75 Resp: 7910
Ion Ratio Lower Upper
75 100
53 119.1 94.0 141.0
89 47.1 37.0 55.4



#82
4-Chlorotoluene
Concen: 19.688 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion: 91 Resp: 84207
Ion Ratio Lower Upper
91 100
126 28.6 14.1 42.4









#83
tert-Butylbenzene
Concen: 20.321 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:119 Resp: 84163
Ion Ratio Lower Upper
119 100
91 61.3 31.4 94.0
134 24.0 11.8 35.3



#84
1,2,4-Trimethylbenzene
Concen: 19.992 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:105 Resp: 88713
Ion Ratio Lower Upper
105 100
120 43.9 21.9 65.5









#85
 sec-Butylbenzene
 Concen: 20.264 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBSD01


Tgt Ion:105 Resp: 108557
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 20.307 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX026587.D
 Acq: 26 Jan 2022 12:24

Tgt Ion:119 Resp: 91269
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.0 39.0
 91 24.7 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 19.925 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01





Tgt Ion:146 Resp: 46044

Ion Ratio Lower Upper

146 100

111 40.9 21.3 63.7

148 62.9 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 19.727 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24




Tgt Ion:146 Resp: 46142


Ion Ratio Lower Upper

146 100

111 40.9 20.2 60.6

148 64.2 31.4 94.0





#89

n-Butylbenzene

Concen: 19.950 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX026587.D

Acq: 26 Jan 2022 12:24

Instrument :

MSVOA_X

ClientSampleId :

VX0127WBSD01





Tgt Ion: 91 Resp: 76947

Ion Ratio Lower Upper

91 100

92 54.1 27.2 81.6

134 26.8 13.1 39.1



#90

Hexachloroethane

Concen: 18.986 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX026587.D




Acq: 26 Jan 2022 12:24


Tgt Ion: 117 Resp: 13181

Ion Ratio Lower Upper

117 100

201 74.9 35.6 106.8









#91
1,2-Dichlorobenzene
Concen: 20.065 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24




Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01


Tgt Ion:146 Resp: 43819
Ion Ratio Lower Upper
146 100
111 42.6 21.8 65.3
148 64.7 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.129 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24


Tgt Ion: 75 Resp: 6329
Ion Ratio Lower Upper
75 100
155 83.2 40.0 119.9
157 108.6 51.6 154.8








#93
1,2,4-Trichlorobenzene
Concen: 19.311 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24


Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01





Tgt Ion:180 Resp: 27037
Ion Ratio Lower Upper
180 100
182 95.3 46.7 140.0
145 32.3 16.3 48.8




#94
Hexachlorobutadiene
Concen: 19.189 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24



Tgt Ion:225 Resp: 11341
Ion Ratio Lower Upper
225 100
223 65.1 32.0 96.0
227 65.6 32.3 96.9









#95
Naphthalene
Concen: 19.380 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Instrument :
MSVOA_X
ClientSampleId :
VX0127WBSD01

Tgt Ion:128 Resp: 90507

Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.3	15.5
129	10.7	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 19.762 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX026587.D
Acq: 26 Jan 2022 12:24

Tgt Ion:180 Resp: 27211

Ion	Ratio	Lower	Upper
180	100		
182	94.4	47.2	141.6
145	33.8	17.5	52.5

