

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026584.D
 Acq On : 26 Jan 2022 11:07
 Operator : JC/MD
 Sample : VX0127MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127MBL01

Quant Time: Jan 26 11:32:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	127503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215072	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82763	46.467	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.940%
35) Dibromofluoromethane	5.385	113	67192	48.379	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	8.646	98	252837	50.518	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.040%
62) 4-Bromofluorobenzene	11.079	95	96196	54.106	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026584.D
Acq On : 26 Jan 2022 11:07
Operator : JC/MD
Sample : VX0127MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0127MBL01

Quant Time: Jan 26 11:32:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

