

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026585.D
 Acq On : 26 Jan 2022 11:30
 Operator : JC/MD
 Sample : VX0127WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127WBL01

Quant Time: Jan 26 11:56:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137743	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232610	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90073	46.811	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.620%
35) Dibromofluoromethane	5.391	113	72538	48.290	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	274650	50.739	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.480%
62) 4-Bromofluorobenzene	11.079	95	96130	49.992	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026585.D
Acq On : 26 Jan 2022 11:30
Operator : JC/MD
Sample : VX0127WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0127WBL01

Quant Time: Jan 26 11:56:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

