

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026590.D
 Acq On : 26 Jan 2022 13:33
 Operator : JC/MD
 Sample : PB142265TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142265TB

Quant Time: Jan 26 13:56:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222689	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204324	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88606	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85425	46.086	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.180%
35) Dibromofluoromethane	5.391	113	69358	48.230	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.460%
50) Toluene-d8	8.653	98	264096	50.963	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.920%
62) 4-Bromofluorobenzene	11.079	95	99174	53.873	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.740%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12287	12.999	ug/l	96
20) Methylene Chloride	2.794	84	4307	2.758	ug/l #	88
43) Isopropyl Acetate	6.348	43	13358	3.601	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026590.D
Acq On : 26 Jan 2022 13:33
Operator : JC/MD
Sample : PB142265TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142265TB

Quant Time: Jan 26 13:56:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

