

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026593.D
 Acq On : 26 Jan 2022 14:42
 Operator : JC/MD
 Sample : PB142265ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142265ZHE#03

Quant Time: Jan 27 07:40:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221255	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85942	46.804	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.600%
35) Dibromofluoromethane	5.391	113	68852	48.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.653	98	262456	50.975	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	97567	53.343	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.680%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5894	6.294	ug/l	99
20) Methylene Chloride	2.794	84	2799	1.809	ug/l	87
43) Isopropyl Acetate	6.348	43	12380	3.359	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026593.D
Acq On : 26 Jan 2022 14:42
Operator : JC/MD
Sample : PB142265ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142265ZHE#03

Quant Time: Jan 27 07:40:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

