

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026594.D
 Acq On : 26 Jan 2022 15:06
 Operator : JC/MD
 Sample : PB142265ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142265ZHE#04

Quant Time: Jan 26 15:35:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86230	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89394	46.876	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.760%
35) Dibromofluoromethane	5.391	113	70686	48.056	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	8.653	98	270089	50.955	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.920%
62) 4-Bromofluorobenzene	11.079	95	98215	52.160	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.320%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6250	6.427	ug/l	90
20) Methylene Chloride	2.794	84	2641	1.644	ug/l	92
43) Isopropyl Acetate	6.342	43	12790	3.371	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026594.D
Acq On : 26 Jan 2022 15:06
Operator : JC/MD
Sample : PB142265ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142265ZHE#04

Quant Time: Jan 26 15:35:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

