

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026597.D
 Acq On : 26 Jan 2022 16:15
 Operator : JC/MD
 Sample : PB142265ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142265ZHE#07

Quant Time: Jan 26 16:39:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233464	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90322	46.659	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.320%
35) Dibromofluoromethane	5.391	113	71989	47.750	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.653	98	274704	50.563	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.079	95	99626	51.621	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.240%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	12952	13.120	ug/l	96
20) Methylene Chloride	2.794	84	4483	2.749	ug/l	96
43) Isopropyl Acetate	6.342	43	14740	3.790	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026597.D
Acq On : 26 Jan 2022 16:15
Operator : JC/MD
Sample : PB142265ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142265ZHE#07

Quant Time: Jan 26 16:39:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

