

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026599.D
 Acq On : 26 Jan 2022 17:01
 Operator : JC/MD
 Sample : PB142265ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142265ZHE#09

Quant Time: Jan 27 07:32:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86271	47.083	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.160%
35) Dibromofluoromethane	5.391	113	68748	48.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.480%
50) Toluene-d8	8.653	98	260790	50.778	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.079	95	94179	51.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.240%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13156	14.079	ug/l	94
20) Methylene Chloride	2.794	84	4276	2.770	ug/l	93
43) Isopropyl Acetate	6.348	43	14520	3.949	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026599.D
Acq On : 26 Jan 2022 17:01
Operator : JC/MD
Sample : PB142265ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142265ZHE#09

Quant Time: Jan 27 07:32:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

