

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026603.D
 Acq On : 26 Jan 2022 18:33
 Operator : JC/MD
 Sample : N1286-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17

Quant Time: Jan 27 07:33:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89915	46.845	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.700%
35) Dibromofluoromethane	5.391	113	71929	48.626	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.653	98	271014	50.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.680%
62) 4-Bromofluorobenzene	11.079	95	96946	51.196	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.400%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	7504	7.666	ug/l	93
37) Ethyl Acetate	4.721	43	30729	13.009	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026603.D
Acq On : 26 Jan 2022 18:33
Operator : JC/MD
Sample : N1286-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17

Quant Time: Jan 27 07:33:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

