

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026606.D
 Acq On : 26 Jan 2022 19:43
 Operator : JC/MD
 Sample : N1286-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8A

Quant Time: Jan 27 07:33:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221156	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87436	47.872	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.740%
35) Dibromofluoromethane	5.391	113	69770	48.853	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	8.653	98	255081	49.564	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.120%
62) 4-Bromofluorobenzene	11.079	95	88592	48.458	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	13202	14.174	ug/l	99
37) Ethyl Acetate	4.721	43	12245	5.369	ug/l	99
59) 2-Hexanone	9.439	43	837	0.500	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026606.D
Acq On : 26 Jan 2022 19:43
Operator : JC/MD
Sample : N1286-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8A

Quant Time: Jan 27 07:33:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

