

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026607.D
 Acq On : 26 Jan 2022 20:06
 Operator : JC/MD
 Sample : N1286-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-20

Quant Time: Jan 27 07:33:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	212868	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83505	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83883	48.206	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.420%
35) Dibromofluoromethane	5.391	113	66659	48.492	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.980%
50) Toluene-d8	8.653	98	251562	50.784	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.079	95	94022	53.431	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.860%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5242	5.907	ug/l	97
37) Ethyl Acetate	4.721	43	29522	13.449	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026607.D
Acq On : 26 Jan 2022 20:06
Operator : JC/MD
Sample : N1286-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-20

Quant Time: Jan 27 07:33:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

