

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026613.D
 Acq On : 27 Jan 2022 10:47
 Operator : JC/MD
 Sample : VX0127MBL02
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0127MBL02

Quant Time: Jan 28 04:23:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88930	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84815	46.697	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.400%
35) Dibromofluoromethane	5.385	113	68943	48.317	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.640%
50) Toluene-d8	8.647	98	257348	50.049	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.079	95	97693	53.483	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026613.D
Acq On : 27 Jan 2022 10:47
Operator : JC/MD
Sample : VX0127MBL02
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0127MBL02

Quant Time: Jan 28 04:23:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

