

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026616.D
 Acq On : 27 Jan 2022 12:04
 Operator : JC/MD
 Sample : N1286-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12R

Quant Time: Jan 28 04:24:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221920	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86039	46.944	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.880%
35) Dibromofluoromethane	5.385	113	69550	48.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.060%
50) Toluene-d8	8.647	98	262633	50.856	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.720%
62) 4-Bromofluorobenzene	11.079	95	98610	53.752	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.500%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	7114	7.611	ug/l	100
37) Ethyl Acetate	4.715	43	20526	8.969	ug/l	99
40) Benzene	6.044	78	2633	0.421	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026616.D
Acq On : 27 Jan 2022 12:04
Operator : JC/MD
Sample : N1286-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12R

Quant Time: Jan 28 04:24:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

