

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026617.D
 Acq On : 27 Jan 2022 12:27
 Operator : JC/MD
 Sample : N1286-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-01-26

Quant Time: Jan 28 04:24:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89691	46.793	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.580%
35) Dibromofluoromethane	5.385	113	72301	48.052	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.653	98	273880	50.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.020%
62) 4-Bromofluorobenzene	11.079	95	99451	51.633	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.260%
Target Compounds						
16) Acetone	2.398	43	2355	2.409	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026617.D
Acq On : 27 Jan 2022 12:27
Operator : JC/MD
Sample : N1286-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-01-26

Quant Time: Jan 28 04:24:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

