

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026620.D
 Acq On : 27 Jan 2022 13:36
 Operator : JC/MD
 Sample : N1286-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-30

Quant Time: Jan 28 04:25:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	141808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240948	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92002	46.443	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.880%
35) Dibromofluoromethane	5.391	113	74511	47.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	8.653	98	284140	50.676	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.360%
62) 4-Bromofluorobenzene	11.079	95	102625	51.523	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.040%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	7127	7.055	ug/l #	88
19) Methyl tert-butyl Ether	3.111	73	8979	1.671	ug/l #	1
22) Diisopropyl ether	3.764	45	2433	0.440	ug/l #	82
31) Cyclohexane	5.471	56	2429	0.831	ug/l	87
37) Ethyl Acetate	4.727	43	11509	4.632	ug/l	99
40) Benzene	6.044	78	24978	3.679	ug/l	96
52) Toluene	8.720	92	5822	1.398	ug/l	99
67) Ethyl Benzene	10.195	91	96776	11.677	ug/l	100
68) m/p-Xylenes	10.305	106	2808	0.905	ug/l	95
69) o-Xylene	10.646	106	1457	0.487	ug/l	95
73) Isopropylbenzene	10.963	105	9349	1.262	ug/l	98
78) n-propylbenzene	11.305	91	22625	2.678	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	8421	1.379	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	32336	5.293	ug/l	98
85) sec-Butylbenzene	11.896	105	1488	0.202	ug/l	99
95) Naphthalene	13.780	128	5458	0.849	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026620.D
Acq On : 27 Jan 2022 13:36
Operator : JC/MD
Sample : N1286-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-30

Quant Time: Jan 28 04:25:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

