

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026625.D
 Acq On : 27 Jan 2022 15:31
 Operator : JC/MD
 Sample : N1296-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8BRA

Quant Time: Jan 28 04:26:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93030	46.597	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.200%
35) Dibromofluoromethane	5.391	113	75453	48.472	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.940%
50) Toluene-d8	8.653	98	279824	49.884	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	96885	48.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.240%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2978	2.925	ug/l	96
17) Carbon Disulfide	2.514	76	1240	0.285	ug/l #	81
19) Methyl tert-butyl Ether	3.117	73	1450	0.268	ug/l #	80
30) Chloroform	5.105	83	4496	1.437	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026625.D
Acq On : 27 Jan 2022 15:31
Operator : JC/MD
Sample : N1296-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8BRA

Quant Time: Jan 28 04:26:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

