

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012722\  
 Data File : VX026626.D  
 Acq On : 27 Jan 2022 15:54  
 Operator : JC/MD  
 Sample : N1296-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-8BRB

Quant Time: Jan 28 04:26:33 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011122W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 11 14:45:13 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.562  | 168   | 137730   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 233730   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 203215   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 81883    | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 90227    | 46.896   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 93.800%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 72552    | 48.068   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 96.140%  |
| 50) Toluene-d8              | 8.653  | 98    | 270714   | 49.772   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 99.540%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 94224    | 48.766   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 97.540%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.392  | 43    | 2361     | 2.406    | ug/l # | 84       |
| 19) Methyl tert-butyl Ether | 3.117  | 73    | 1650     | 0.316    | ug/l # | 83       |
| 30) Chloroform              | 5.104  | 83    | 4868     | 1.615    | ug/l   | 99       |
| 52) Toluene                 | 8.720  | 92    | 2977     | 0.737    | ug/l   | 97       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012722\  
Data File : VX026626.D  
Acq On : 27 Jan 2022 15:54  
Operator : JC/MD  
Sample : N1296-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-8BRB

Quant Time: Jan 28 04:26:33 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011122W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jan 11 14:45:13 2022  
Response via : Initial Calibration

