

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026630.D
 Acq On : 27 Jan 2022 17:26
 Operator : JC/MD
 Sample : N1296-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-2A

Quant Time: Jan 28 04:27:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214359	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83923	47.652	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.300%
35) Dibromofluoromethane	5.391	113	68278	49.324	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.640%
50) Toluene-d8	8.653	98	254147	50.949	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.900%
62) 4-Bromofluorobenzene	11.079	95	89532	50.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.060%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1774	1.975	ug/l #	89
20) Methylene Chloride	2.794	84	575	0.387	ug/l #	92
30) Chloroform	5.099	83	9653	3.497	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026630.D
Acq On : 27 Jan 2022 17:26
Operator : JC/MD
Sample : N1296-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-2A

Quant Time: Jan 28 04:27:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

