

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026635.D
 Acq On : 27 Jan 2022 19:21
 Operator : JC/MD
 Sample : N1297-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-1

Quant Time: Jan 28 04:27:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218189	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84749	47.293	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.580%
35) Dibromofluoromethane	5.391	113	67719	48.062	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	8.653	98	255269	50.275	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.079	95	88673	49.162	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.320%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1940	2.123	ug/l	90
30) Chloroform	5.099	83	17166	6.113	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	2559	0.483	ug/l	97
95) Naphthalene	13.780	128	1115	0.200	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026635.D
Acq On : 27 Jan 2022 19:21
Operator : JC/MD
Sample : N1297-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-1

Quant Time: Jan 28 04:27:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

