

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026636.D
 Acq On : 27 Jan 2022 19:44
 Operator : JC/MD
 Sample : N1297-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Jan 28 04:28:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	125569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	215238	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83616	47.669	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.340%
35) Dibromofluoromethane	5.391	113	67999	48.922	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.840%
50) Toluene-d8	8.653	98	250414	49.995	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.000%
62) 4-Bromofluorobenzene	11.079	95	86778	48.771	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.540%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2062	2.305	ug/l	100
30) Chloroform	5.099	83	30184	10.980	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026636.D
Acq On : 27 Jan 2022 19:44
Operator : JC/MD
Sample : N1297-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Jan 28 04:28:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

