

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012722\
 Data File : VX026610.D
 Acq On : 26 Jan 2022 21:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 07:34:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	40.362	19.3	77	0.00
3 P	Chloromethane	50.000	44.482	11.0	88	0.00
4 C	Vinyl Chloride	50.000	44.561	10.9#	84	0.00
5 T	Bromomethane	50.000	40.385	19.2	79	0.00
6 T	Chloroethane	50.000	42.530	14.9	82	0.00
7 T	Trichlorofluoromethane	50.000	42.423	15.2	79	0.00
8 T	Diethyl Ether	50.000	45.600	8.8	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	45.400	9.2	85	0.00
10 T	Methyl Iodide	50.000	49.540	0.9	98	0.00
11 T	Tert butyl alcohol	250.000	254.208	-1.7	93	0.00
12 CM	1,1-Dichloroethene	50.000	45.703	8.6#	85	0.00
13 T	Acrolein	250.000	187.316	25.1#	76	0.00
14 T	Allyl chloride	50.000	46.038	7.9	84	0.00
15 T	Acrylonitrile	250.000	250.905	-0.4	95	0.00
16 T	Acetone	250.000	179.530	28.2#	67	0.00
17 T	Carbon Disulfide	50.000	39.153	21.7	76	0.00
18 T	Methyl Acetate	50.000	50.779	-1.6	98	0.00
19 T	Methyl tert-butyl Ether	50.000	46.436	7.1	87	0.00
20 T	Methylene Chloride	50.000	46.398	7.2	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.334	9.3	85	0.00
22 T	Diisopropyl ether	50.000	47.398	5.2	88	0.00
23 T	Vinyl Acetate	250.000	242.232	3.1	88	0.00
24 P	1,1-Dichloroethane	50.000	46.892	6.2	87	0.00
25 T	2-Butanone	250.000	222.222	11.1	82	0.00
26 T	2,2-Dichloropropane	50.000	44.620	10.8	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.536	4.9	88	0.00
28 T	Bromochloromethane	50.000	46.646	6.7	86	0.00
29 T	Tetrahydrofuran	250.000	250.829	-0.3	95	0.00
30 C	Chloroform	50.000	47.429	5.1#	89	0.00
31 T	Cyclohexane	50.000	43.494	13.0	83	0.00
32 T	1,1,1-Trichloroethane	50.000	46.101	7.8	83	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.148	9.7	84	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	46.880	6.2	87	0.00
36 T	1,1-Dichloropropene	50.000	45.046	9.9	85	0.00
37 T	Ethyl Acetate	50.000	49.568	0.9	92	0.00
38 T	Carbon Tetrachloride	50.000	45.925	8.2	83	0.00
39 T	Methylcyclohexane	50.000	45.000	10.0	84	0.00
40 TM	Benzene	50.000	47.419	5.2	89	0.00
41 T	Methacrylonitrile	50.000	49.553	0.9	92	0.00
42 TM	1,2-Dichloroethane	50.000	45.598	8.8	85	0.00
43 T	Isopropyl Acetate	50.000	49.823	0.4	91	0.00
44 TM	Trichloroethene	50.000	48.497	3.0	90	0.00
45 C	1,2-Dichloropropane	50.000	49.568	0.9#	92	0.00
46 T	Dibromomethane	50.000	48.758	2.5	91	0.00
47 T	Bromodichloromethane	50.000	48.560	2.9	88	0.00
48 T	Methyl methacrylate	50.000	49.584	0.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012722\
 Data File : VX026610.D
 Acq On : 26 Jan 2022 21:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 27 07:34:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	935.309	6.5	85	0.00
50 S	Toluene-d8	50.000	48.165	3.7	90	0.00
51 T	4-Methyl-2-Pentanone	250.000	250.573	-0.2	93	0.00
52 CM	Toluene	50.000	49.062	1.9#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	46.202	7.6	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.623	-1.2	89	0.00
55 T	1,1,2-Trichloroethane	50.000	49.994	0.0	93	0.00
56 T	Ethyl methacrylate	50.000	51.698	-3.4	92	0.00
57 T	1,3-Dichloropropane	50.000	50.219	-0.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	241.938	3.2	89	0.00
59 T	2-Hexanone	250.000	236.252	5.5	86	0.00
60 T	Dibromochloromethane	50.000	52.002	-4.0	93	0.00
61 T	1,2-Dibromoethane	50.000	51.000	-2.0	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.597	2.8	90	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	49.923	0.2	98	0.00
65 PM	Chlorobenzene	50.000	48.104	3.8	95	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.889	0.2	96	0.00
67 C	Ethyl Benzene	50.000	47.790	4.4#	91	0.00
68 T	m/p-Xylenes	100.000	96.406	3.6	92	0.00
69 T	o-Xylene	50.000	48.335	3.3	92	0.00
70 T	Styrene	50.000	49.467	1.1	92	0.00
71 P	Bromoform	50.000	45.300	9.4	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	48.221	3.6	92	0.00
74 T	N-amyl acetate	50.000	50.471	-0.9	91	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.266	5.5	91	0.00
76 T	1,2,3-Trichloropropane	50.000	48.621	2.8	91	0.00
77 T	Bromobenzene	50.000	47.984	4.0	92	0.00
78 T	n-propylbenzene	50.000	48.262	3.5	91	0.00
79 T	2-Chlorotoluene	50.000	46.873	6.3	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.108	3.8	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.341	7.3	94	0.00
82 T	4-Chlorotoluene	50.000	46.654	6.7	90	0.00
83 T	tert-Butylbenzene	50.000	48.228	3.5	91	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.985	4.0	91	0.00
85 T	sec-Butylbenzene	50.000	48.648	2.7	91	0.00
86 T	p-Isopropyltoluene	50.000	48.603	2.8	91	0.00
87 T	1,3-Dichlorobenzene	50.000	47.185	5.6	92	0.00
88 T	1,4-Dichlorobenzene	50.000	46.655	6.7	90	0.00
89 T	n-Butylbenzene	50.000	48.070	3.9	89	0.00
90 T	Hexachloroethane	50.000	44.820	10.4	90	0.00
91 T	1,2-Dichlorobenzene	50.000	47.920	4.2	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.719	0.6	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.542	4.9	89	0.00
94 T	Hexachlorobutadiene	50.000	46.134	7.7	88	0.00
95 T	Naphthalene	50.000	49.758	0.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026610.D
Acq On : 26 Jan 2022 21:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 27 07:34:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.250	3.5	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6