

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026612.D
 Acq On : 27 Jan 2022 10:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 28 04:23:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.590	0.518	12.2	83	0.00
3 P	Chloromethane	0.541	0.526	2.8	94	0.00
4 C	Vinyl Chloride	0.568	0.541	4.8#	88	0.00
5 T	Bromomethane	0.262	0.269	-2.7	99	0.00
6 T	Chloroethane	0.287	0.259	9.8	86	0.00
7 T	Trichlorofluoromethane	0.897	0.821	8.5	84	0.00
8 T	Diethyl Ether	0.317	0.304	4.1	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.538	1.6	91	0.00
10 T	Methyl Iodide	0.535	0.684	-27.9#	114	0.00
11 T	Tert butyl alcohol	0.097	0.081	16.5	75	0.03
12 CM	1,1-Dichloroethene	0.522	0.501	4.0#	88	0.00
13 T	Acrolein	0.103	0.073	29.1#	71	0.00
14 T	Allyl chloride	0.960	0.924	3.7	87	0.00
15 T	Acrylonitrile	0.299	0.302	-1.0	94	0.00
16 T	Acetone	0.356	0.333	6.5	86	0.00
17 T	Carbon Disulfide	1.525	1.308	14.2	81	0.00
18 T	Methyl Acetate	0.708	0.706	0.3	95	0.00
19 T	Methyl tert-butyl Ether	1.895	1.798	5.1	88	0.00
20 T	Methylene Chloride	0.589	0.573	2.7	95	0.00
21 T	trans-1,2-Dichloroethene	0.565	0.539	4.6	88	0.00
22 T	Diisopropyl ether	1.951	1.928	1.2	91	0.00
23 T	Vinyl Acetate	1.606	1.569	2.3	87	0.00
24 P	1,1-Dichloroethane	1.053	1.029	2.3	90	0.00
25 T	2-Butanone	0.463	0.462	0.2	90	0.00
26 T	2,2-Dichloropropane	0.781	0.782	-0.1	88	0.00
27 T	cis-1,2-Dichloroethene	0.645	0.637	1.2	90	0.00
28 T	Bromochloromethane	0.433	0.413	4.6	86	0.00
29 T	Tetrahydrofuran	0.275	0.266	3.3	90	0.00
30 C	Chloroform	1.095	1.065	2.7#	90	0.00
31 T	Cyclohexane	1.030	0.954	7.4	87	0.00
32 T	1,1,1-Trichloroethane	0.963	0.935	2.9	87	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.588	15.8	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.323	0.294	9.0	82	0.00
36 T	1,1-Dichloropropene	0.510	0.497	2.5	89	0.00
37 T	Ethyl Acetate	0.516	0.511	1.0	90	0.00
38 T	Carbon Tetrachloride	0.524	0.521	0.6	87	0.00
39 T	Methylcyclohexane	0.615	0.614	0.2	90	0.00
40 TM	Benzene	1.409	1.411	-0.1	92	0.00
41 T	Methacrylonitrile	0.292	0.304	-4.1	94	0.00
42 TM	1,2-Dichloroethane	0.551	0.529	4.0	87	0.00
43 T	Isopropyl Acetate	0.833	0.842	-1.1	89	0.00
44 TM	Trichloroethene	0.400	0.420	-5.0	94	0.00
45 C	1,2-Dichloropropane	0.352	0.367	-4.3#	94	0.00
46 T	Dibromomethane	0.248	0.255	-2.8	93	0.00
47 T	Bromodichloromethane	0.501	0.524	-4.6	92	0.00
48 T	Methyl methacrylate	0.413	0.422	-2.2	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
 Data File : VX026612.D
 Acq On : 27 Jan 2022 10:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 28 04:23:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.005	28.6#	67	0.00
50 S	Toluene-d8	1.164	1.114	4.3	87	0.00
51 T	4-Methyl-2-Pentanone	0.494	0.507	-2.6	92	0.00
52 CM	Toluene	0.864	0.933	-8.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.470	0.545	-16.0	97	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.589	-9.5	93	0.00
55 T	1,1,2-Trichloroethane	0.337	0.370	-9.8	99	0.00
56 T	Ethyl methacrylate	0.517	0.577	-11.6	97	0.00
57 T	1,3-Dichloropropane	0.585	0.636	-8.7	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.271	0.190	29.9#	63	0.00
59 T	2-Hexanone	0.378	0.391	-3.4	91	0.00
60 T	Dibromochloromethane	0.351	0.412	-17.4	101	0.00
61 T	1,2-Dibromoethane	0.351	0.387	-10.3	99	0.00
62 S	4-Bromofluorobenzene	0.413	0.419	-1.5	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.472	0.506	-7.2	105	0.00
65 PM	Chlorobenzene	1.053	1.094	-3.9	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.413	-5.6	101	0.00
67 C	Ethyl Benzene	1.913	1.998	-4.4#	100	0.00
68 T	m/p-Xylenes	0.716	0.767	-7.1	102	0.00
69 T	o-Xylene	0.691	0.743	-7.5	103	0.00
70 T	Styrene	1.127	1.230	-9.1	102	0.00
71 P	Bromoform	0.267	0.300	-12.4	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.098	4.359	-6.4	103	0.00
74 T	N-amyl acetate	1.582	1.590	-0.5	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.100	1.072	2.5	95	0.00
76 T	1,2,3-Trichloropropane	1.101	1.117	-1.5	97	0.00
77 T	Bromobenzene	0.939	0.990	-5.4	102	0.00
78 T	n-propylbenzene	4.674	4.989	-6.7	102	0.00
79 T	2-Chlorotoluene	2.834	2.906	-2.5	101	0.00
80 T	1,3,5-Trimethylbenzene	3.379	3.570	-5.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.312	0.350	-12.2	104	0.00
82 T	4-Chlorotoluene	3.257	3.346	-2.7	100	0.00
83 T	tert-Butylbenzene	3.154	3.276	-3.9	99	0.00
84 T	1,2,4-Trimethylbenzene	3.380	3.550	-5.0	101	0.00
85 T	sec-Butylbenzene	4.080	4.404	-7.9	102	0.00
86 T	p-Isopropyltoluene	3.423	3.703	-8.2	102	0.00
87 T	1,3-Dichlorobenzene	1.760	1.819	-3.4	102	0.00
88 T	1,4-Dichlorobenzene	1.781	1.842	-3.4	101	0.00
89 T	n-Butylbenzene	2.937	3.253	-10.8	104	0.00
90 T	Hexachloroethane	0.509	0.576	-13.2	103	0.00
91 T	1,2-Dichlorobenzene	1.663	1.750	-5.2	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.266	0.251	5.6	87	0.00
93 T	1,2,4-Trichlorobenzene	1.066	1.132	-6.2	101	0.00
94 T	Hexachlorobutadiene	0.450	0.488	-8.4	105	0.00
95 T	Naphthalene	3.557	3.699	-4.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012722\
Data File : VX026612.D
Acq On : 27 Jan 2022 10:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 28 04:23:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.049	1.132	-7.9	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6