

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012722\  
 Data File : VX026615.D  
 Acq On : 27 Jan 2022 11:37  
 Operator : JC/MD  
 Sample : VX0127WBS02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0127WBS02

Quant Time: Jan 28 04:24:13 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011122W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 11 14:45:13 2022  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
 Supervised By :Semsettin Yesilyurt 02/09/2022

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 108766              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 177183              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 159511              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 72405               | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 74621               | 49.113  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 98.220%  |         |        |          |
| 35) Dibromofluoromethane     | 5.385          | 113  | 59380               | 51.897  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 103.800% |         |        |          |
| 50) Toluene-d8               | 8.647          | 98   | 223113              | 54.112  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 108.220% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 83045               | 56.697  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 113.400% |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.173          | 85   | 21843               | 17.007  | ug/l   | 100      |
| 3) Chloromethane             | 1.295          | 50   | 21886               | 18.589  | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.380          | 62   | 22951               | 18.566  | ug/l   | 99       |
| 5) Bromomethane              | 1.618          | 94   | 13020               | 22.832  | ug/l   | 97       |
| 6) Chloroethane              | 1.697          | 64   | 11961               | 19.135  | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 1.892          | 101  | 34561               | 17.709  | ug/l   | 100      |
| 8) Diethyl Ether             | 2.136          | 74   | 12997               | 18.819  | ug/l   | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 2.337          | 101  | 22626               | 19.024  | ug/l   | 98       |
| 10) Methyl Iodide            | 2.459          | 142  | 26942               | 23.617  | ug/l   | 98       |
| 11) Tert butyl alcohol       | 3.020          | 59   | 20252               | 95.817  | ug/l # | 75       |
| 12) 1,1-Dichloroethene       | 2.325          | 96   | 21255               | 18.712  | ug/l   | 97       |
| 13) Acrolein                 | 2.246          | 56   | 17739               | 78.982  | ug/l   | 99       |
| 14) Allyl chloride           | 2.666          | 41   | 38613               | 18.484  | ug/l   | 98       |
| 15) Acrylonitrile            | 3.069          | 53   | 65302               | 100.367 | ug/l   | 99       |
| 16) Acetone                  | 2.392          | 43   | 70872               | 91.468  | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.520          | 76   | 53999               | 16.282  | ug/l   | 100      |
| 18) Methyl Acetate           | 2.709          | 43   | 31096               | 20.187  | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 3.111          | 73   | 77281               | 18.750  | ug/l   | 99       |
| 20) Methylene Chloride       | 2.794          | 84   | 24383               | 19.047  | ug/l   | 98       |
| 21) trans-1,2-Dichloroethene | 3.099          | 96   | 23398               | 19.027  | ug/l   | 95       |
| 22) Diisopropyl ether        | 3.758          | 45   | 83028               | 19.567  | ug/l # | 86       |
| 23) Vinyl Acetate            | 3.721          | 43   | 333977              | 95.617  | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 43403               | 18.940  | ug/l   | 100      |
| 25) 2-Butanone               | 4.562          | 43   | 98129               | 97.385  | ug/l   | 100      |
| 26) 2,2-Dichloropropane      | 4.483          | 77   | 31080               | 18.297  | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.495          | 96   | 27617               | 19.678  | ug/l   | 98       |
| 28) Bromochloromethane       | 4.898          | 49   | 19966               | 21.218  | ug/l   | 100      |
| 29) Tetrahydrofuran          | 5.013          | 42   | 57049               | 95.327  | ug/l   | 97       |
| 30) Chloroform               | 5.099          | 83   | 45913               | 19.283  | ug/l   | 99       |
| 31) Cyclohexane              | 5.477          | 56   | 41378               | 18.464  | ug/l   | 99       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 39227               | 18.730  | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 33556               | 18.577  | ug/l   | 98       |
| 37) Ethyl Acetate            | 4.715          | 43   | 36052               | 19.731  | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 34856               | 18.760  | ug/l   | 99       |
| 39) Methylcyclohexane        | 7.379          | 83   | 41448               | 19.024  | ug/l   | 97       |
| 40) Benzene                  | 6.038          | 78   | 97535               | 19.536  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012722\  
 Data File : VX026615.D  
 Acq On : 27 Jan 2022 11:37  
 Operator : JC/MD  
 Sample : VX0127WBS02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0127WBS02

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
 Supervised By :Semsettin Yesilyurt 02/09/2022

Quant Time: Jan 28 04:24:13 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011122W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 11 14:45:13 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.922  | 41   | 20400    | 19.709  | ug/l   | 97       |
| 42) 1,2-Dichloroethane        | 6.086  | 62   | 36561    | 18.717  | ug/l   | 100      |
| 43) Isopropyl Acetate         | 6.342  | 43   | 58190    | 19.715  | ug/l   | 99       |
| 44) Trichloroethene           | 7.129  | 130  | 28420    | 20.061  | ug/l   | 96       |
| 45) 1,2-Dichloropropane       | 7.428  | 63   | 25081    | 20.126  | ug/l   | 99       |
| 46) Dibromomethane            | 7.580  | 93   | 17358    | 19.759  | ug/l   | 99       |
| 47) Bromodichloromethane      | 7.824  | 83   | 35251    | 19.860  | ug/l   | 98       |
| 48) Methyl methacrylate       | 7.696  | 41   | 29255    | 20.000  | ug/l   | 98       |
| 49) 1,4-Dioxane               | 7.702  | 88   | 8776     | 365.603 | ug/l   | 98       |
| 51) 4-Methyl-2-Pentanone      | 8.574  | 43   | 183360   | 104.738 | ug/l   | 99       |
| 52) Toluene                   | 8.720  | 92   | 63297    | 20.675  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 8.976  | 75   | 34269    | 19.261  | ug/l   | 96       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 38886    | 20.397  | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 25776    | 21.574  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 39262    | 21.443  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 43691    | 21.093  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.245  | 63   | 107001   | 111.449 | ug/l   | 100      |
| 59) 2-Hexanone                | 9.433  | 43   | 138642   | 103.472 | ug/l   | 99       |
| 60) Dibromochloromethane      | 9.519  | 129  | 26732    | 21.471  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 26424    | 21.225  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 30609    | 20.314  | ug/l   | 97       |
| 65) Chlorobenzene             | 10.080 | 112  | 67103    | 19.977  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 24861    | 19.919  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 123773   | 20.277  | ug/l   | 100      |
| 68) m/p-Xylenes               | 10.299 | 106  | 93059    | 40.737  | ug/l   | 99       |
| 69) o-Xylene                  | 10.640 | 106  | 45279    | 20.542  | ug/l   | 99       |
| 70) Styrene                   | 10.659 | 104  | 75694    | 21.045  | ug/l   | 99       |
| 71) Bromoform                 | 10.799 | 173  | 17167    | 19.410  | ug/l # | 99       |
| 73) Isopropylbenzene          | 10.964 | 105  | 119945   | 20.210  | ug/l   | 98       |
| 74) N-amyl acetate            | 10.842 | 43   | 44612    | 19.471  | ug/l   | 97       |
| 75) 1,1,2,2-Tetrachloroethane | 11.214 | 83   | 31184    | 19.576  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.238 | 75   | 32443m   | 20.351  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 27554    | 20.266  | ug/l   | 98       |
| 78) n-propylbenzene           | 11.305 | 91   | 139125   | 20.557  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 82206    | 20.030  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 101742   | 20.794  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 8847     | 20.738  | ug/l   | 99       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 94673    | 20.071  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.713 | 119  | 92512    | 20.254  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 100243   | 20.483  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.890 | 105  | 122373   | 20.712  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 103559   | 20.892  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 51578    | 20.238  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 51797    | 20.080  | ug/l   | 98       |
| 89) n-Butylbenzene            | 12.335 | 91   | 89012    | 20.926  | ug/l   | 100      |
| 90) Hexachloroethane          | 12.543 | 117  | 15170    | 19.718  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 48691    | 20.217  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 7030     | 18.259  | ug/l   | 95       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 30513    | 19.761  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 13130    | 20.144  | ug/l   | 100      |
| 95) Naphthalene               | 13.774 | 128  | 102144   | 19.832  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 30310    | 19.960  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012722\  
Data File : VX026615.D  
Acq On : 27 Jan 2022 11:37  
Operator : JC/MD  
Sample : VX0127WBS02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02

Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022

Quant Time: Jan 28 04:24:13 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011122W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jan 11 14:45:13 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:168 Resp: 108760  
Ion Ratio Lower Upper  
168 100  
99 59.8 50.3 75.5

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#2  
Dichlorodifluoromethane  
Concen: 17.007 ug/l  
RT: 1.173 min Scan# 14  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 85 Resp: 21843  
Ion Ratio Lower Upper  
85 100  
87 32.5 16.4 49.1



Abundance Scan 34 (1.294 min): VX026417.D\data.ms (-29)



#3

Chloromethane

Concen: 18.589 ug/l

RT: 1.295 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 50 Resp: 2188

Ion Ratio Lower Upper

50 100





52 32.2 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022

Supervised By :Semsettin Yesilyurt 02/09/2022



#4

Vinyl Chloride

Concen: 18.566 ug/l

RT: 1.380 min Scan# 48

Delta R.T. 0.000 min

Lab File: VX026615.D




Acq: 27 Jan 2022 11:37

Tgt Ion: 62 Resp: 22951


Ion Ratio Lower Upper

62 100

64 31.9 26.2 39.2



Abundance Scan 87 (1.617 min): VX026417.D\data.ms (-82)



#5

Bromomethane

Concen: 22.832 ug/l

RT: 1.618 min Scan# 81

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 94 Resp: 13020

Ion Ratio Lower Upper

94 100


96 98.2 75.9 113.9

Manual Integrations


APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022


Abundance Scan 87 (1.618 min): VX026615.D\data.ms




Abundance Scan 87 (1.618 min): VX026615.D\data.ms (-38)



Abundance



Abundance Scan 99 (1.691 min): VX026417.D\data.ms (-93)



#6

Chloroethane

Concen: 19.135 ug/l

RT: 1.697 min Scan# 100

Delta R.T. 0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37


Tgt Ion: 64 Resp: 11961

Ion Ratio Lower Upper


64 100

66 33.7 25.4 38.2


Abundance Scan 100 (1.697 min): VX026615.D\data.ms




Abundance Scan 100 (1.697 min): VX026615.D\data.ms (-50)




Abundance





#7  
Trichlorofluoromethane  
Concen: 17.709 ug/l  
RT: 1.892 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:101 Resp: 3456  
Ion Ratio Lower Upper  
101 100  
103 64.3 51.3 76.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#8  
Diethyl Ether  
Concen: 18.819 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 74 Resp: 12997  
Ion Ratio Lower Upper  
74 100  
45 99.6 50.7 152.1





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 19.024 ug/l  
RT: 2.337 min Scan# 205  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02



Tgt Ion:101 Resp: 22620  
Ion Ratio Lower Upper  
101 100  
85 43.5 35.3 52.9  
151 78.6 61.4 92.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#10  
Methyl Iodide  
Concen: 23.617 ug/l  
RT: 2.459 min Scan# 225  
Delta R.T. -0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion:142 Resp: 26942  
Ion Ratio Lower Upper  
142 100  
127 46.2 38.1 57.1  
141 14.3 11.2 16.8





#11  
Tert butyl alcohol  
Concen: 95.817 ug/l  
RT: 3.020 min Scan# 315  
Delta R.T. 0.013 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 59 Resp: 2025  
Ion Ratio Lower Upper  
59 100  
57 0.0 7.2 10.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#12  
1,1-Dichloroethene  
Concen: 18.712 ug/l  
RT: 2.325 min Scan# 203  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 96 Resp: 21255  
Ion Ratio Lower Upper  
96 100  
61 168.5 138.5 207.7  
98 63.4 51.8 77.6



Abundance Scan 189 (2.239 min): VX026417.D\data.ms (-18



#13

Acrolein

Concen: 78.982 ug/l

RT: 2.246 min Scan# 190

Delta R.T. 0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 56 Resp: 17739

Ion Ratio Lower Upper

56 100

55 70.2 56.9 85.3


Manual Integrations

APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022

Supervised By :Semsettin Yesilyurt 02/09/2022


Abundance Scan 190 (2.246 min): VX026615.D\data.ms




Abundance



Abundance Scan 190 (2.246 min): VX026615.D\data.ms (-14



Abundance Scan 260 (2.672 min): VX026417.D\data.ms (-25



#14

Allyl chloride

Concen: 18.484 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Tgt Ion: 41 Resp: 38613


Ion Ratio Lower Upper

41 100


39 65.0 54.2 81.4

76 31.7 25.2 37.8


Abundance Scan 259 (2.666 min): VX026615.D\data.ms




Abundance




Abundance Scan 259 (2.666 min): VX026615.D\data.ms (-21





#15  
Acrylonitrile  
Concen: 100.367 ug/l  
RT: 3.069 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02



Tgt Ion: 53 Resp: 65302  
Ion Ratio Lower Upper  
53 100  
52 82.8 66.6 100.0  
51 36.1 29.5 44.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#16  
Acetone  
Concen: 91.468 ug/l  
RT: 2.392 min Scan# 214  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion: 43 Resp: 70872  
Ion Ratio Lower Upper  
43 100  
58 29.8 23.0 34.6





#17  
Carbon Disulfide  
Concen: 16.282 ug/l  
RT: 2.520 min Scan# 211  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 76 Resp: 5399  
Ion Ratio Lower Upper  
76 100  
78 9.5 7.5 11.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#18  
Methyl Acetate  
Concen: 20.187 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 43 Resp: 31096  
Ion Ratio Lower Upper  
43 100  
74 23.6 18.3 27.5









#19  
Methyl tert-butyl Ether  
Concen: 18.750 ug/l  
RT: 3.111 min Scan# 311  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02

Tgt Ion: 73 Resp: 7728  
Ion Ratio Lower Upper  
73 100  
57 22.2 17.4 26.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#20  
Methylene Chloride  
Concen: 19.047 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion: 84 Resp: 24383  
Ion Ratio Lower Upper  
84 100  
49 127.5 101.5 152.3  
51 41.4 31.2 46.8  
86 66.1 51.3 76.9





#21  
trans-1,2-Dichloroethene  
Concen: 19.027 ug/l  
RT: 3.099 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 96 Resp: 23398  
Ion Ratio Lower Upper  
96 100  
61 134.0 112.9 169.3  
98 60.4 49.5 74.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#22  
Diisopropyl ether  
Concen: 19.567 ug/l  
RT: 3.758 min Scan# 438  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 45 Resp: 83028  
Ion Ratio Lower Upper  
45 100  
43 93.4 60.6 91.0#  
87 26.2 22.0 33.0  
59 10.8 9.0 13.4





#23  
 Vinyl Acetate  
 Concen: 95.617 ug/l  
 RT: 3.721 min Scan# 41  
 Delta R.T. -0.006 min  
 Lab File: VX026615.D  
 Acq: 27 Jan 2022 11:37




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0127WBS02




Tgt Ion: 43 Resp: 33397  
 Ion Ratio Lower Upper  
 43 100  
 86 10.4 8.1 12.1

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
 Supervised By :Semsettin Yesilyurt 02/09/2022




#24  
 1,1-Dichloroethane  
 Concen: 18.940 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.006 min  
 Lab File: VX026615.D  
 Acq: 27 Jan 2022 11:37



Tgt Ion: 63 Resp: 43403  
 Ion Ratio Lower Upper  
 63 100  
 98 7.2 3.7 11.1  
 100 4.8 2.4 7.2





#25

2-Butanone

Concen: 97.385 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :





VX0127WBS02

Tgt Ion: 43 Resp: 98129  
 Ion Ratio Lower Upper  
 43 100  
 72 25.7 20.4 30.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
 Supervised By :Semsettin Yesilyurt 02/09/2022



#26

2,2-Dichloropropane

Concen: 18.297 ug/l




RT: 4.483 min Scan# 557


Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37


Tgt Ion: 77 Resp: 31080  
 Ion Ratio Lower Upper  
 77 100  
 97 22.9 11.0 33.0





#27  
cis-1,2-Dichloroethene  
Concen: 19.678 ug/l  
RT: 4.495 min Scan# 51  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 96 Resp: 2761  
Ion Ratio Lower Upper  
96 100  
61 145.0 0.0 296.6  
98 63.9 0.0 128.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#28  
Bromochloromethane  
Concen: 21.218 ug/l  
RT: 4.898 min Scan# 625  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 49 Resp: 19966  
Ion Ratio Lower Upper  
49 100  
129 1.8 0.0 4.2  
130 78.3 62.4 93.6





#29  
Tetrahydrofuran  
Concen: 95.327 ug/l  
RT: 5.013 min Scan# 644  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 42 Resp: 57049  
Ion Ratio Lower Upper  
42 100  
72 44.8 34.6 52.0  
71 41.6 31.8 47.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#30  
Chloroform  
Concen: 19.283 ug/l  
RT: 5.099 min Scan# 658  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 83 Resp: 45913  
Ion Ratio Lower Upper  
83 100  
85 63.8 51.6 77.4





#31  
Cyclohexane  
Concen: 18.464 ug/l  
RT: 5.477 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 56 Resp: 41378  
Ion Ratio Lower Upper  
56 100  
69 30.7 25.8 38.6  
84 86.7 68.9 103.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#32  
1,1,1-Trichloroethane  
Concen: 18.730 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 97 Resp: 39227  
Ion Ratio Lower Upper  
97 100  
99 63.9 51.3 76.9  
61 45.7 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 49.113 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 65 Resp: 7462

Ion Ratio Lower Upper

65 100





67 51.4 0.0 100.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022

Supervised By :Semsettin Yesilyurt 02/09/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37




Tgt Ion:114 Resp: 177183


Ion Ratio Lower Upper

114 100

63 20.2 0.0 42.8


88 15.5 0.0 33.0





#35  
Dibromofluoromethane  
Concen: 51.897 ug/l  
RT: 5.385 min Scan# 706  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 113 Resp: 59380  
Ion Ratio Lower Upper  
113 100  
111 103.1 81.2 121.8  
192 18.5 15.0 22.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#36  
1,1-Dichloropropene  
Concen: 18.577 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 75 Resp: 33556  
Ion Ratio Lower Upper  
75 100  
110 35.8 17.2 51.6  
77 30.2 24.6 36.8









#37  
Ethyl Acetate  
Concen: 19.731 ug/l  
RT: 4.715 min Scan# 595  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02

Tgt Ion: 43 Resp: 36052  
Ion Ratio Lower Upper  
43 100  
61 13.4 10.6 16.0  
70 11.0 8.5 12.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#38  
Carbon Tetrachloride  
Concen: 18.760 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Tgt Ion:117 Resp: 34856  
Ion Ratio Lower Upper  
117 100  
119 93.2 74.7 112.1  
121 29.9 25.4 38.0





#39

Methylcyclohexane

Concen: 19.024 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02





Tgt Ion: 83 Resp: 4144

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 55  | 76.7  | 63.6  | 95.4  |
| 98  | 49.4  | 38.4  | 57.6  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#40

Benzene

Concen: 19.536 ug/l

RT: 6.038 min Scan# 812




Delta R.T. -0.006 min


Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Tgt Ion: 78 Resp: 97535


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 78  | 100   |       |       |
| 77  | 22.3  | 18.9  | 28.3  |





#41  
Methacrylonitrile  
Concen: 19.709 ug/l  
RT: 4.922 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02



Tgt Ion: 41 Resp: 20400  
Ion Ratio Lower Upper  
41 100  
39 55.7 45.4 68.0  
67 71.4 53.2 79.8  
52 29.5 23.0 34.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#42  
1,2-Dichloroethane  
Concen: 18.717 ug/l  
RT: 6.086 min Scan# 820  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion: 62 Resp: 36561  
Ion Ratio Lower Upper  
62 100  
98 8.9 0.0 17.8





#43  
Isopropyl Acetate  
Concen: 19.715 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 43 Resp: 58190  
Ion Ratio Lower Upper  
43 100  
61 19.5 15.4 23.0  
87 13.2 10.2 15.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#44  
Trichloroethene  
Concen: 20.061 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion:130 Resp: 28420  
Ion Ratio Lower Upper  
130 100  
95 96.8 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.126 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 63 Resp: 2508

Ion Ratio Lower Upper

63 100





65 30.1 23.8 35.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022

Supervised By :Semsettin Yesilyurt 02/09/2022



#46

Dibromomethane

Concen: 19.759 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37




Tgt Ion: 93 Resp: 17358


Ion Ratio Lower Upper

93 100

95 83.2 66.1 99.1


174 99.8 78.7 118.1





#47  
Bromodichloromethane  
Concen: 19.860 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 83 Resp: 3525  
Ion Ratio Lower Upper  
83 100  
85 62.3 51.4 77.0  
127 8.6 6.9 10.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#48  
Methyl methacrylate  
Concen: 20.000 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 41 Resp: 29255  
Ion Ratio Lower Upper  
41 100  
69 78.4 62.2 93.4  
39 54.3 45.4 68.0





#49  
1,4-Dioxane  
Concen: 365.603 ug/l  
RT: 7.702 min Scan# 1085  
Delta R.T. 0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02



Tgt Ion: 88 Resp: 8770  
Ion Ratio Lower Upper  
88 100  
43 35.0 29.9 44.9  
58 71.8 57.8 86.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#50  
Toluene-d8  
Concen: 54.112 ug/l  
RT: 8.647 min Scan# 1240  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion: 98 Resp: 223113  
Ion Ratio Lower Upper  
98 100  
100 66.8 53.2 79.8





#51  
4-Methyl-2-Pentanone  
Concen: 104.738 ug/l  
RT: 8.574 min Scan# 1228  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 43 Resp: 183360  
Ion Ratio Lower Upper  
43 100  
58 38.5 30.5 45.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#52  
Toluene  
Concen: 20.675 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 92 Resp: 63297  
Ion Ratio Lower Upper  
92 100  
91 170.2 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 19.261 ug/l

RT: 8.976 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 75 Resp: 34269

Ion Ratio Lower Upper





75 100

77 34.6 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#54

cis-1,3-Dichloropropene

Concen: 20.397 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37




Tgt Ion: 75 Resp: 38886


Ion Ratio Lower Upper

75 100

77 32.9 26.2 39.2


39 50.9 43.5 65.3





#55  
1,1,2-Trichloroethane  
Concen: 21.574 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02






Tgt Ion: 97 Resp: 25770


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 84.3  | 68.4  | 102.6 |
| 85  | 53.9  | 44.0  | 66.0  |
| 99  | 64.5  | 49.8  | 74.6  |

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022






#56  
Ethyl methacrylate  
Concen: 21.443 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 69 Resp: 39262


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 68.2  | 57.2  | 85.8  |
| 39  | 37.0  | 30.9  | 46.3  |





#57  
1,3-Dichloropropane  
Concen: 21.093 ug/l  
RT: 9.311 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 76 Resp: 4369  
Ion Ratio Lower Upper  
76 100  
78 32.2 25.7 38.5

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#58  
2-Chloroethyl Vinyl ether  
Concen: 111.449 ug/l  
RT: 8.245 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 63 Resp: 107001  
Ion Ratio Lower Upper  
63 100  
106 28.8 23.0 34.6





#59

2-Hexanone

Concen: 103.472 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :

VX0127WBS02

Tgt Ion: 43 Resp: 13864

Ion Ratio Lower Upper





43 100

58 53.3 26.4 79.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#60

Dibromochloromethane

Concen: 21.471 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.006 min

Lab File: VX026615.D




Acq: 27 Jan 2022 11:37


Tgt Ion:129 Resp: 26732

Ion Ratio Lower Upper

129 100

127 77.0 38.5 115.5





#61

1,2-Dibromoethane

Concen: 21.225 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37

Instrument :

MSVOA\_X

ClientSampleId :





VX0127WBS02

Tgt Ion:107 Resp: 26424  
 Ion Ratio Lower Upper  
 107 100  
 109 93.1 75.5 113.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
 Supervised By :Semsettin Yesilyurt 02/09/2022



#62

4-Bromofluorobenzene

Concen: 56.697 ug/l




RT: 11.079 min Scan# 1639


Delta R.T. 0.000 min

Lab File: VX026615.D

Acq: 27 Jan 2022 11:37


Tgt Ion: 95 Resp: 83045  
 Ion Ratio Lower Upper  
 95 100  
 174 74.6 0.0 148.6  
 176 72.0 0.0 141.4





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02



Tgt Ion:117 Resp: 15951  
Ion Ratio Lower Upper  
117 100  
82 56.2 45.2 67.8  
119 32.2 26.2 39.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#64  
Tetrachloroethene  
Concen: 20.314 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion:164 Resp: 30609  
Ion Ratio Lower Upper  
164 100  
166 129.6 101.8 152.6  
129 97.4 83.3 124.9  
131 97.9 78.0 117.0





#65  
Chlorobenzene  
Concen: 19.977 ug/l  
RT: 10.080 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:112 Resp: 6710  
Ion Ratio Lower Upper  
112 100  
114 31.9 25.3 37.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 19.919 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion:131 Resp: 24861  
Ion Ratio Lower Upper  
131 100  
133 95.9 47.9 143.8  
119 64.9 31.7 95.1









#67  
Ethyl Benzene  
Concen: 20.277 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02

Tgt Ion: 91 Resp: 12377  
Ion Ratio Lower Upper  
91 100  
106 31.1 24.7 37.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#68  
m/p-Xylenes  
Concen: 40.737 ug/l  
RT: 10.299 min Scan# 1511  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion:106 Resp: 93059  
Ion Ratio Lower Upper  
106 100  
91 205.4 165.3 247.9





#69  
o-Xylene  
Concen: 20.542 ug/l  
RT: 10.640 min Scan# 1567  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:106 Resp: 45279  
Ion Ratio Lower Upper  
106 100  
91 218.1 109.7 329.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#70  
Styrene  
Concen: 21.045 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion:104 Resp: 75694  
Ion Ratio Lower Upper  
104 100  
78 51.4 42.5 63.7  
103 54.2 43.3 64.9





#71  
Bromoform  
Concen: 19.410 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. -0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:173 Resp: 1716  
Ion Ratio Lower Upper  
173 100  
175 49.0 24.3 72.9  
254 0.0 0.0 0.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion:152 Resp: 72405  
Ion Ratio Lower Upper  
152 100  
115 72.1 44.2 132.6  
150 166.2 0.0 348.2





#73  
Isopropylbenzene  
Concen: 20.210 ug/l  
RT: 10.964 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:105 Resp: 11994  
Ion Ratio Lower Upper  
105 100  
120 26.9 13.1 39.1

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#74  
N-ethyl acetate  
Concen: 19.471 ug/l  
RT: 10.842 min Scan# 1600  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 43 Resp: 44612  
Ion Ratio Lower Upper  
43 100  
70 42.7 32.9 49.3  
55 25.2 18.3 27.5  
61 24.9 18.9 28.3





#75  
1,1,2,2-Tetrachloroethane  
Concen: 19.576 ug/l  
RT: 11.214 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 83 Resp: 31184  
Ion Ratio Lower Upper  
83 100  
131 10.9 5.3 15.9  
85 64.1 32.0 96.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#76  
1,2,3-Trichloropropane  
Concen: 20.351 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. -0.006 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 75 Resp: 32443  
Ion Ratio Lower Upper  
75 100  
77 37.7 18.3 54.8





#77  
Bromobenzene  
Concen: 20.266 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:156 Resp: 27554  
Ion Ratio Lower Upper  
156 100  
77 165.9 84.5 253.4  
158 97.2 48.1 144.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#78  
n-propylbenzene  
Concen: 20.557 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion: 91 Resp: 139125  
Ion Ratio Lower Upper  
91 100  
120 23.2 11.6 34.7









#79  
2-Chlorotoluene  
Concen: 20.030 ug/l  
RT: 11.366 min Scan# 11366  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02

Tgt Ion: 91 Resp: 82200  
Ion Ratio Lower Upper  
91 100  
126 33.2 16.4 49.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#80  
1,3,5-Trimethylbenzene  
Concen: 20.794 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion:105 Resp: 101742  
Ion Ratio Lower Upper  
105 100  
120 47.9 24.0 72.0





#81  
trans-1,4-Dichloro-2-butene  
Concen: 20.738 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 75 Resp: 884  
Ion Ratio Lower Upper  
75 100  
53 117.0 94.0 141.0  
89 47.0 37.0 55.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#82  
4-Chlorotoluene  
Concen: 20.071 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 91 Resp: 94673  
Ion Ratio Lower Upper  
91 100  
126 28.8 14.1 42.4





#83  
tert-Butylbenzene  
Concen: 20.254 ug/l  
RT: 11.713 min Scan# 1743  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:119 Resp: 92512  
Ion Ratio Lower Upper  
119 100  
91 62.8 31.4 94.0  
134 24.5 11.8 35.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#84  
1,2,4-Trimethylbenzene  
Concen: 20.483 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion:105 Resp: 100243  
Ion Ratio Lower Upper  
105 100  
120 44.3 21.9 65.5





#85  
 sec-Butylbenzene  
 Concen: 20.712 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. 0.000 min  
 Lab File: VX026615.D  
 Acq: 27 Jan 2022 11:37




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0127WBS02




Tgt Ion:105 Resp: 12237  
 Ion Ratio Lower Upper  
 105 100  
 134 20.5 10.1 30.1

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
 Supervised By :Semsettin Yesilyurt 02/09/2022




#86  
 p-Isopropyltoluene  
 Concen: 20.892 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. 0.000 min  
 Lab File: VX026615.D  
 Acq: 27 Jan 2022 11:37



Tgt Ion:119 Resp: 103559  
 Ion Ratio Lower Upper  
 119 100  
 134 26.4 13.0 39.0  
 91 24.4 12.6 37.6









#87  
1,3-Dichlorobenzene  
Concen: 20.238 ug/l  
RT: 11.969 min Scan# 1785  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02

Tgt Ion:146 Resp: 51578  
Ion Ratio Lower Upper  
146 100  
111 41.4 21.3 63.7  
148 62.3 31.6 95.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#88  
1,4-Dichlorobenzene  
Concen: 20.080 ug/l  
RT: 12.043 min Scan# 1797  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37


Tgt Ion:146 Resp: 51797  
Ion Ratio Lower Upper  
146 100  
111 41.4 20.2 60.6  
148 64.5 31.4 94.0





#89  
n-Butylbenzene  
Concen: 20.926 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion: 91 Resp: 8901  
Ion Ratio Lower Upper  
91 100  
92 54.2 27.2 81.6  
134 26.0 13.1 39.1

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#90  
Hexachloroethane  
Concen: 19.718 ug/l  
RT: 12.543 min Scan# 1879  
Delta R.T. 0.001 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion:117 Resp: 15170  
Ion Ratio Lower Upper  
117 100  
201 70.8 35.6 106.8





#91  
1,2-Dichlorobenzene  
Concen: 20.217 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:146 Resp: 4869  
Ion Ratio Lower Upper  
146 100  
111 43.4 21.8 65.3  
148 63.5 32.0 96.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 18.259 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion: 75 Resp: 7030  
Ion Ratio Lower Upper  
75 100  
155 85.0 40.0 119.9  
157 107.0 51.6 154.8





#93  
1,2,4-Trichlorobenzene  
Concen: 19.761 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:180 Resp: 3051  
Ion Ratio Lower Upper  
180 100  
182 94.5 46.7 140.0  
145 32.3 16.3 48.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022




#94  
Hexachlorobutadiene  
Concen: 20.144 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Tgt Ion:225 Resp: 13130  
Ion Ratio Lower Upper  
225 100  
223 64.2 32.0 96.0  
227 64.4 32.3 96.9





#95  
Naphthalene  
Concen: 19.832 ug/l  
RT: 13.774 min Scan# 2081  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0127WBS02




Tgt Ion:128 Resp: 102144  
Ion Ratio Lower Upper  
128 100  
127 12.8 10.3 15.5  
129 10.8 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 02/09/2022  
Supervised By :Semsettin Yesilyurt 02/09/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 19.960 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.000 min  
Lab File: VX026615.D  
Acq: 27 Jan 2022 11:37



Tgt Ion:180 Resp: 30310  
Ion Ratio Lower Upper  
180 100  
182 94.7 47.2 141.6  
145 33.7 17.5 52.5

