

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012723\
 Data File : VX033929.D
 Acq On : 27 Jan 2023 11:55
 Operator : JC/MD
 Sample : 01323-16DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22DL

Quant Time: Jan 28 03:03:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	166189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101352	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78446	44.831	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.660%
35) Dibromofluoromethane	5.385	113	72858	49.251	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.500%
50) Toluene-d8	8.647	98	267166	49.567	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.140%
62) 4-Bromofluorobenzene	11.079	95	92100	46.022	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.040%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	26073	18.520	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	72727	40.016	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012723\
Data File : VX033929.D
Acq On : 27 Jan 2023 11:55
Operator : JC/MD
Sample : 01323-16DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22DL

Quant Time: Jan 28 03:03:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

