



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012723\
Data File : VX033922.D
Acq On : 27 Jan 2023 08:25
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Title : SW846 8260
Last Update : Wed Jan 11 14:48:42 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	3363	PASS
75	95	30	60	47.6	9809	PASS
95	95	100	100	100.0	20605	PASS
96	95	5	9	6.7	1388	PASS
173	174	0.00	2	0.7	129	PASS
174	95	50	100	93.2	19207	PASS
175	174	5	9	7.6	1461	PASS
176	174	95	101	97.7	18771	PASS
177	176	5	9	6.7	1263	PASS