

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
 Data File : VX044718.D
 Acq On : 27 Jan 2025 13:01
 Operator : JC/MD
 Sample : Q1173-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-60-62

Quant Time: Jan 28 04:11:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 16 01:16:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	145080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	291470	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	260564	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	111378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	107332	49.013	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	98.020%	
35) Dibromofluoromethane	5.379	113	95202	51.414	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.820%	
50) Toluene-d8	8.647	98	359433	55.671	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	111.340%	
62) 4-Bromofluorobenzene	11.079	95	125581	52.176	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	104.360%	
Target Compounds						
16) Acetone	2.386	43	10408	12.013	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012725\
Data File : VX044718.D
Acq On : 27 Jan 2025 13:01
Operator : JC/MD
Sample : Q1173-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-60-62

Quant Time: Jan 28 04:11:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011525W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 16 01:16:09 2025
Response via : Initial Calibration

